

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050523\
 Data File : VD076020.D
 Acq On : 05 May 2023 12:36
 Operator : KP/SY
 Sample : 02589-05
 Misc : 4.93G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JREB9

Quant Time: May 06 00:22:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 06 00:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	496934	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	396398	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	110063	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	152874	14.023	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.080%
7) Chloroethane-d5	2.799	69	155648	16.525	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.080%
11) 1,1-Dichloroethene-d2	3.922	65	37108	14.765	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.040%
21) 2-Butanone-d5	6.993	46	48752	51.514	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.020%
24) Chloroform-d	7.569	84	249530	20.080	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.320%
26) 1,2-Dichloroethane-d4	8.228	65	108159	20.382	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.520%
32) Benzene-d6	8.204	84	448302	19.404	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.600%
36) 1,2-Dichloropropane-d6	9.210	67	154894	24.221	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.880%
41) Toluene-d8	10.269	98	376867	17.424	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.680%
43) trans-1,3-Dichloroprop...	10.528	79	51660	19.544	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.160%
47) 2-Hexanone-d5	10.881	63	38126	49.515	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.020%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	116848	23.878	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.520%
66) 1,2-Dichlorobenzene-d4	13.816	152	92660	23.801	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.200%
Target Compounds						Qvalue
16) Methylene chloride	4.799	84	12328	1.223	ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050523\
Data File : VD076020.D
Acq On : 05 May 2023 12:36
Operator : KP/SY
Sample : 02589-05
Misc : 4.93G/10.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JREB9

Quant Time: May 06 00:22:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 06 00:21:26 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 1.223 ug/L
RT: 4.799 min Scan# 54
Delta R.T. 0.000 min
Lab File: VD076020.D
Acq: 05 May 2023 12:36

Instrument :
MSVOA_D
ClientSampleId :
JREB9

Tgt Ion: 84 Resp: 12328
Ion Ratio Lower Upper
84 100
86 62.1 45.8 85.2
49 100.5 60.8 112.8

