

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD050619\
 Data File : VD062130.D
 Acq On : 6 May 2019 14:56
 Operator : FY/SY
 Sample : K2668-05RE
 Misc : 4.75g/5mL/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3RE

Quant Time: May 07 05:39:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 02 07:55:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	359053	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.22	114	600919	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	581618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	236422	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	182175	64.83	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	129.66%	
35) Dibromofluoromethane	6.92	113	250377	60.25	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	120.50%	
50) Toluene-d8	10.40	98	523145	52.90	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	105.80%	
62) 4-Bromofluorobenzene	13.56	95	244615	61.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.34%	
Target Compounds						
16) Acetone	3.21	43	51628	76.944	ug/l #	90
20) Methylene Chloride	3.81	84	34216	7.782	ug/l	91
84) 1,2,4-Trimethylbenzene	14.23	105	30734	2.560	ug/l	94
95) Naphthalene	16.12	128	12828	2.253	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD050619\
Data File : VD062130.D
Acq On : 6 May 2019 14:56
Operator : FY/SY
Sample : K2668-05RE
Misc : 4.75g/5mL/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3RE

Quant Time: May 07 05:39:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Thu May 02 07:55:52 2019
Response via : Initial Calibration

