

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069079.D
 Acq On : 06 May 2021 16:56
 Operator : VA/SY
 Sample : M2296-01
 Misc : 7.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1

Quant Time: May 07 01:11:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	34486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	57187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	72029	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	37243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	24810	55.82	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.64%
35) Dibromofluoromethane	7.908	113	22835	58.23	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.46%
50) Toluene-d8	10.338	98	66618	46.01	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.02%
62) 4-Bromofluorobenzene	12.626	95	29640	62.24	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	124.48%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069079.D
Acq On : 06 May 2021 16:56
Operator : VA/SY
Sample : M2296-01
Misc : 7.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1

Quant Time: May 07 01:11:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

