

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069082.D
 Acq On : 06 May 2021 18:25
 Operator : VA/SY
 Sample : M2266-01
 Misc : 5.43G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CRT-016

Quant Time: May 07 01:11:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	250291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	447233	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	513723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	263664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	141262	43.79	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.58%
35) Dibromofluoromethane	7.914	113	157878	51.48	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.96%
50) Toluene-d8	10.338	98	579245	51.16	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.32%
62) 4-Bromofluorobenzene	12.632	95	244091	65.54	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069082.D
Acq On : 06 May 2021 18:25
Operator : VA/SY
Sample : M2266-01
Misc : 5.43G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CRT-016

Quant Time: May 07 01:11:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

