

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069086.D
 Acq On : 06 May 2021 20:24
 Operator : VA/SY
 Sample : M2279-02
 Misc : 5.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EME-13

Quant Time: May 07 01:12:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	246437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	465740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	492834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	236328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	157114	49.47	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.94%		
35) Dibromofluoromethane	7.914	113	166597	52.16	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	104.32%		
50) Toluene-d8	10.338	98	587938	49.86	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	99.72%		
62) 4-Bromofluorobenzene	12.626	95	227965	58.78	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	117.56%		
Target Compounds						Qvalue
17) Carbon Disulfide	4.403	76	9364	1.07	ug/l #	92
25) 2-Butanone	7.220	43	3972	5.85	ug/l	97
30) Chloroform	7.709	83	173093	29.50	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069086.D
Acq On : 06 May 2021 20:24
Operator : VA/SY
Sample : M2279-02
Misc : 5.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EME-13

Quant Time: May 07 01:12:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

