

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069088.D
 Acq On : 06 May 2021 21:23
 Operator : VA/SY
 Sample : M2279-04
 Misc : 6.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EME-15

Quant Time: May 07 01:13:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	247267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	460128	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	491751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	227134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	159414	50.02	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 100.04%		
35) Dibromofluoromethane	7.914	113	166668	52.82	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 105.64%		
50) Toluene-d8	10.338	98	595700	51.14	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 102.28%		
62) 4-Bromofluorobenzene	12.632	95	223141	58.24	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 116.48%		
Target Compounds						
30) Chloroform	7.709	83	294112	49.95	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069088.D
Acq On : 06 May 2021 21:23
Operator : VA/SY
Sample : M2279-04
Misc : 6.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EME-15

Quant Time: May 07 01:13:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

