

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069089.D
 Acq On : 06 May 2021 21:52
 Operator : VA/SY
 Sample : M2279-05
 Misc : 5.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EME-16

Quant Time: May 07 01:13:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	159162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	306144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	337755	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	159934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	121933	59.44	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 118.88%		
35) Dibromofluoromethane	7.908	113	118778	56.58	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 113.16%		
50) Toluene-d8	10.338	98	396685	51.18	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 102.36%		
62) 4-Bromofluorobenzene	12.632	95	148578	58.28	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 116.56%		
Target Compounds						Qvalue
30) Chloroform	7.708	83	132866	35.06	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069089.D
Acq On : 06 May 2021 21:52
Operator : VA/SY
Sample : M2279-05
Misc : 5.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EME-16

Quant Time: May 07 01:13:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

