

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069092.D
 Acq On : 06 May 2021 23:19
 Operator : VA/SY
 Sample : M2279-08
 Misc : 5.95G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EME-19

Quant Time: May 07 01:15:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	68623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	122359	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	145182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	67442	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	59742	67.55	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 135.10%			
35) Dibromofluoromethane	7.920	113	51569	61.46	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 122.92%			
50) Toluene-d8	10.338	98	157798	50.94	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 101.88%			
62) 4-Bromofluorobenzene	12.626	95	57633	56.56	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 113.12%			
Target Compounds						Qvalue
16) Acetone	4.150	43	5188	18.88	ug/l #	81
30) Chloroform	7.708	83	38719	23.70	ug/l	97
84) 1,2,4-Trimethylbenzene	13.261	105	4036	1.02	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069092.D
Acq On : 06 May 2021 23:19
Operator : VA/SY
Sample : M2279-08
Misc : 5.95G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EME-19

Quant Time: May 07 01:15:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

