

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
 Data File : VD069076.D
 Acq On : 06 May 2021 15:00
 Operator : VA/SY
 Sample : VD0506SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0506SBL01

Quant Time: May 07 01:09:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	297834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	514857	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	579899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	306496	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	146180	38.08	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.16%
35) Dibromofluoromethane	7.914	113	167046	47.31	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.62%
50) Toluene-d8	10.338	98	632258	48.51	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.02%
62) 4-Bromofluorobenzene	12.632	95	268212	62.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.12%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050621\
Data File : VD069076.D
Acq On : 06 May 2021 15:00
Operator : VA/SY
Sample : VD0506SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0506SBL01

Quant Time: May 07 01:09:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00




Instrument :
MSVOA_D
ClientSampleId :
VD0506SBL01


Tgt Ion:168 Resp: 297834
Ion Ratio Lower Upper
168 100
99 55.5 52.6 78.8



#33
1,2-Dichloroethane-d4
Concen: 38.08 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.006 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Tgt Ion: 65 Resp: 146180
Ion Ratio Lower Upper
65 100
67 53.3 0.0 103.4









#34

1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Instrument :
MSVOA_D
ClientSampleId :
VD0506SBL01




Tgt Ion:114 Resp: 514857
Ion Ratio Lower Upper
114 100
63 18.6 0.0 41.4
88 15.3 0.0 27.8




#35

Dibromofluoromethane
Concen: 47.31 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. -0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Tgt Ion:113 Resp: 167046
Ion Ratio Lower Upper
113 100
111 101.5 84.6 126.8
192 19.3 13.5 20.3









#50
Toluene-d8
Concen: 48.51 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00




Instrument :
MSVOA_D
ClientSampleId :
VD0506SBL01


Tgt Ion: 98 Resp: 632258
Ion Ratio Lower Upper
98 100
100 65.0 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 62.56 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Tgt Ion: 95 Resp: 268212
Ion Ratio Lower Upper
95 100
174 85.1 0.0 156.0
176 81.2 0.0 149.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.644 min Scan# 1653
Delta R.T. -0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Instrument :
MSVOA_D
ClientSampleId :
VD0506SBL01

Tgt Ion:117 Resp: 579899
Ion Ratio Lower Upper
117 100
82 51.9 45.4 68.2
119 31.6 26.8 40.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. -0.000 min
Lab File: VD069076.D
Acq: 06 May 2021 15:00

Tgt Ion:152 Resp: 306496
Ion Ratio Lower Upper
152 100
115 57.4 30.5 91.5
150 158.0 0.0 340.8

