

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050724\
 Data File : VD078882.D
 Acq On : 07 May 2024 16:57
 Operator : RP/MD
 Sample : P2395-02
 Misc : 5.06G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-DRIVEWAY

Quant Time: May 08 04:23:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050124S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 02 07:55:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	29869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	73850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	84307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	37183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	29920	90.247	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	180.500%#
35) Dibromofluoromethane	7.804	113	33037	60.928	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.860%
50) Toluene-d8	10.269	98	97620	46.912	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.569	95	27172	45.625	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.260%
Target Compounds						
20) Methylene Chloride	4.793	84	3820	7.772	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050724\
Data File : VD078882.D
Acq On : 07 May 2024 16:57
Operator : RP/MD
Sample : P2395-02
Misc : 5.06G/5ml/MSVOA_D/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-DRIVEWAY

Quant Time: May 08 04:23:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050124S.M
Quant Title : SW846 8260
QLast Update : Thu May 02 07:55:11 2024
Response via : Initial Calibration

