

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD050819\
 Data File : VD062174.D
 Acq On : 8 May 2019 15:15
 Operator : FY/SY
 Sample : K2717-02
 Misc : 6.38g/5mL/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P026-SS001-0206-01

Quant Time: May 09 01:24:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	15767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.19	114	24154	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	20265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	9487	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.42	65	10569	72.13	ug/l	0.00
Spiked Amount			Recovery	=	144.26%	
35) Dibromofluoromethane	6.89	113	10351	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
50) Toluene-d8	10.39	98	26042	53.89	ug/l	0.00
Spiked Amount			Recovery	=	107.78%	
62) 4-Bromofluorobenzene	13.55	95	10531	54.27	ug/l	0.00
Spiked Amount			Recovery	=	108.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.13	94	2834	112.944	ug/l #	12
10) Methyl Iodide	3.32	142	391	1.643	ug/l #	37
20) Methylene Chloride	3.81	84	1039	8.098	ug/l #	8
38) Carbon Tetrachloride	7.03	117	931	3.397	ug/l #	12
74) N-amyl acetate	13.69	43	245	1.643	ug/l #	49
81) trans-1,4-Dichloro-2-buten	13.55	75	4232	132.962	ug/l #	16
95) Naphthalene	16.11	128	260	1.068	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD050819\
Data File : VD062174.D
Acq On : 8 May 2019 15:15
Operator : FY/SY
Sample : K2717-02
Misc : 6.38g/5mL/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P026-SS001-0206-01

Quant Time: May 09 01:24:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration

