

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD050819\
 Data File : VD062178.D
 Acq On : 8 May 2019 17:06
 Operator : FY/SY
 Sample : K2672-19
 Misc : 4.99g/5mL/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-1.0-1.5

Quant Time: May 09 01:25:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	3783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.19	114	5305	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.35	117	3577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	1510	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.42	65	2837	80.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	161.40%	
35) Dibromofluoromethane	6.89	113	2032	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
50) Toluene-d8	10.39	98	5285	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	13.54	95	1571	36.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.72%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.18	94	248	41.193	ug/l #	1
10) Methyl Iodide	3.30	142	184	3.222	ug/l #	37
23) Vinyl Acetate	5.08	43	227	4.821	ug/l #	77
25) 2-Butanone	6.13	43	305	37.927	ug/l #	56
37) Ethyl Acetate	6.13	43	305	16.535	ug/l #	64
74) N-amyl acetate	13.29	43	220	9.271	ug/l #	49
81) trans-1,4-Dichloro-2-buten	13.54	75	587	115.870	ug/l #	16

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD050819\
Data File : VD062178.D
Acq On : 8 May 2019 17:06
Operator : FY/SY
Sample : K2672-19
Misc : 4.99g/5mL/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-1.0-1.5

Quant Time: May 09 01:25:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration

