

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD050819\
 Data File : VD062179.D
 Acq On : 8 May 2019 17:34
 Operator : FY/SY
 Sample : K2717-07
 Misc : 4.12g/5mL/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P026-SS002-0206-01

Quant Time: May 09 01:25:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	12982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	27036	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.35	117	29980	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	12318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	26801	222.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	444.30%	
35) Dibromofluoromethane	6.91	113	20817	92.80	ug/l	0.02
Spiked Amount 50.000			Recovery	=	185.60%	
50) Toluene-d8	10.39	98	32072	59.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.60%	
62) 4-Bromofluorobenzene	13.54	95	13521	62.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.52%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.19	94	1128	54.598	ug/l #	12
10) Methyl Iodide	3.31	142	842	4.296	ug/l #	37
16) Acetone	3.21	43	282	11.146	ug/l #	50
20) Methylene Chloride	3.81	84	2315	21.915	ug/l #	87
36) 1,1-Dichloropropene	7.04	75	537	2.423	ug/l #	40
38) Carbon Tetrachloride	7.03	117	656	2.138	ug/l #	12
81) trans-1,4-Dichloro-2-buten	13.54	75	5267	127.448	ug/l #	16

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD050819\
Data File : VD062179.D
Acq On : 8 May 2019 17:34
Operator : FY/SY
Sample : K2717-07
Misc : 4.12g/5mL/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P026-SS002-0206-01

Quant Time: May 09 01:25:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration

