

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD050922\  
 Data File : VD072892.D  
 Acq On : 09 May 2022 15:56  
 Operator : VA/SY  
 Sample : N2673-01  
 Misc : 5.03G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 SWEEPINGS-050322

Quant Time: May 09 17:28:05 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D050222S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 03 08:33:49 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.973	168	50639	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	90952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	110812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	53649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	47318	71.985	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.960%
35) Dibromofluoromethane	7.908	113	38034	59.543	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.080%
50) Toluene-d8	10.332	98	115783	50.531	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.060%
62) 4-Bromofluorobenzene	12.620	95	52407	61.340	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.680%
Target Compounds						
52) Toluene	10.396	92	15506	8.287	ug/l	99
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD050922\  
 Data File : VD072892.D  
 Acq On : 09 May 2022 15:56  
 Operator : VA/SY  
 Sample : N2673-01  
 Misc : 5.03G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 SWEEPINGS-050322

Quant Time: May 09 17:28:05 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D050222S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 03 08:33:49 2022  
 Response via : Initial Calibration

