

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051021\
 Data File : VD069118.D
 Acq On : 10 May 2021 12:54
 Operator : VA/SY
 Sample : VD0510SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0510SBL01

Quant Time: May 10 16:10:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	259250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	464147	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	445191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	193605	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	154904	46.36	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.72%
35) Dibromofluoromethane	7.914	113	161177	50.64	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.28%
50) Toluene-d8	10.338	98	577092	49.11	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.22%
62) 4-Bromofluorobenzene	12.626	95	176851	45.76	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.52%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051021\
Data File : VD069118.D
Acq On : 10 May 2021 12:54
Operator : VA/SY
Sample : VD0510SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0510SBL01

Quant Time: May 10 16:10:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

