

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051021\
 Data File : VD069128.D
 Acq On : 10 May 2021 17:39
 Operator : VA/SY
 Sample : M2325-02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: May 11 01:04:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	226557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	424527	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	410650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	177992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	138598	47.47	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.94%
35) Dibromofluoromethane	7.914	113	151278	51.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.92%
50) Toluene-d8	10.338	98	532362	49.53	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.06%
62) 4-Bromofluorobenzene	12.632	95	161148	45.58	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.16%
Target Compounds						
64) Tetrachloroethene	10.867	164	2991	1.13	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051021\
Data File : VD069128.D
Acq On : 10 May 2021 17:39
Operator : VA/SY
Sample : M2325-02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: May 11 01:04:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

