



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051024\
Data File : VD078895.D
Acq On : 10 May 2024 08:46
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051024S.M
Title : SW846 8260
Last Update : Sat May 11 05:14:02 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	11615	PASS
75	95	30	60	50.1	34605	PASS
95	95	100	100	100.0	69051	PASS
96	95	5	9	6.4	4402	PASS
173	174	0.00	2	0.9	558	PASS
174	95	50	100	94.5	65248	PASS
175	174	5	9	8.1	5306	PASS
176	174	95	101	96.3	62843	PASS
177	176	5	9	7.1	4457	PASS