

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051222\
 Data File : VD072963.D
 Acq On : 12 May 2022 22:40
 Operator : VA/SY
 Sample : N2817-05
 Misc : 5.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS021-0006-01

Quant Time: May 13 05:59:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	302807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	547499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	539234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	216230	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	177955	45.273	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.540%
35) Dibromofluoromethane	7.908	113	186443	48.488	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.980%
50) Toluene-d8	10.331	98	673239	48.810	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.620	95	225438	43.834	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051222\
Data File : VD072963.D
Acq On : 12 May 2022 22:40
Operator : VA/SY
Sample : N2817-05
Misc : 5.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS021-0006-01

Quant Time: May 13 05:59:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

