

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051223\
 Data File : VD076102.D
 Acq On : 12 May 2023 12:07
 Operator : KP/SY
 Sample : 02591-36
 Misc : 4.10G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JREA7

Quant Time: May 14 23:55:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM051123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun May 14 23:54:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	396540	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	363529	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	139511	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	155351	25.610	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 102.440%		
7) Chloroethane-d5	2.799	69	142411	25.571	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 102.280%		
11) 1,1-Dichloroethene-d2	3.911	65	37325	24.441	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 97.760%		
21) 2-Butanone-d5	6.987	46	38427	47.049	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 94.100%		
24) Chloroform-d	7.569	84	204639	24.633	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 98.520%		
26) 1,2-Dichloroethane-d4	8.228	65	88747m	24.302	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 97.200%		
32) Benzene-d6	8.199	84	408103	24.901	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 99.600%		
36) 1,2-Dichloropropane-d6	9.210	67	131622	26.063	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 104.240%		
41) Toluene-d8	10.269	98	348716	23.875	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 95.520%		
43) trans-1,3-Dichloroprop...	10.528	79	42597	22.470	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 89.880%		
47) 2-Hexanone-d5	10.881	63	30020	40.593	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 81.180%		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	94960	21.836	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 87.360%		
66) 1,2-Dichlorobenzene-d4	13.816	152	99332	24.924	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 99.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051223\
Data File : VD076102.D
Acq On : 12 May 2023 12:07
Operator : KP/SY
Sample : 02591-36
Misc : 4.10G/10.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JREA7

Quant Time: May 14 23:55:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM051123SMA.M
Quant Title : SFAM01.0
QLast Update : Sun May 14 23:54:23 2023
Response via : Initial Calibration

