

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051321\
 Data File : VD069160.D
 Acq On : 13 May 2021 16:36
 Operator : VA/SY
 Sample : M2372-05
 Misc : 4.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-26598

Quant Time: May 14 01:22:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	8007	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	16246	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	23152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	11418	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	8313	103.56	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 207.12%#	
35) Dibromofluoromethane	7.914	113	5603	53.00	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.00%	
50) Toluene-d8	10.338	98	25358	65.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 130.56%	
62) 4-Bromofluorobenzene	12.626	95	10252	82.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 164.04%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051321\
Data File : VD069160.D
Acq On : 13 May 2021 16:36
Operator : VA/SY
Sample : M2372-05
Misc : 4.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-26598

Quant Time: May 14 01:22:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

