

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051321\
 Data File : VD069147.D
 Acq On : 13 May 2021 10:17
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 14 01:18:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.532	0.517	2.8	95	0.00
3 P	Chloromethane	0.611	0.594	2.8	99	0.00
4 C	Vinyl Chloride	0.805	0.811	-0.7#	95	0.00
5 T	Bromomethane	0.606	0.538	11.2	93	0.00
6 T	Chloroethane	0.518	0.511	1.4	93	0.00
7 T	Trichlorofluoromethane	0.982	0.966	1.6	98	0.00
8 T	Diethyl Ether	0.224	0.207	7.6	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.561	-0.7	98	0.00
10 T	Methyl Iodide	0.531	0.528	0.6	83	0.00
11 T	Tert butyl alcohol	0.034	0.026	23.5	93	-0.01
12 CM	1,1-Dichloroethene	0.500	0.510	-2.0#	94	0.00
13 T	Acrolein	0.041	0.041	0.0	92	0.00
14 T	Allyl chloride	0.549	0.535	2.6	91	0.00
15 T	Acrylonitrile	0.084	0.080	4.8	87	0.00
16 T	Acetone	0.077	0.087	-13.0	102	0.00
17 T	Carbon Disulfide	1.658	1.660	-0.1	97	0.00
18 T	Methyl Acetate	0.187	0.181	3.2	89	0.00
19 T	Methyl tert-butyl Ether	0.908	0.880	3.1	86	0.00
20 T	Methylene Chloride	0.813	0.605	25.6#	86	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.595	-0.3	94	0.00
22 T	Diisopropyl ether	1.096	1.107	-1.0	90	0.00
23 T	Vinyl Acetate	0.647	0.661	-2.2	87	0.00
24 P	1,1-Dichloroethane	0.947	0.935	1.3	94	0.00
25 T	2-Butanone	0.104	0.104	0.0	91	0.00
26 T	2,2-Dichloropropane	0.885	0.903	-2.0	96	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.632	1.9	93	0.00
28 T	Bromochloromethane	0.338	0.319	5.6	90	-0.01
29 T	Tetrahydrofuran	0.057	0.056	1.8	86	0.00
30 C	Chloroform	1.066	1.043	2.2#	93	0.00
31 T	Cyclohexane	0.743	0.730	1.7	98	0.00
32 T	1,1,1-Trichloroethane	0.967	0.975	-0.8	96	0.00
33 S	1,2-Dichloroethane-d4	0.501	0.462	7.8	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.325	0.328	-0.9	97	0.00
36 T	1,1-Dichloropropene	0.486	0.545	-12.1	98	0.00
37 T	Ethyl Acetate	0.156	0.146	6.4	82	0.00
38 T	Carbon Tetrachloride	0.543	0.585	-7.7	97	0.00
39 T	Methylcyclohexane	0.513	0.580	-13.1	96	0.00
40 TM	Benzene	1.392	1.466	-5.3	94	0.00
41 T	Methacrylonitrile	0.083	0.087	-4.8	96	0.00
42 TM	1,2-Dichloroethane	0.380	0.397	-4.5	91	0.00
43 T	Isopropyl Acetate	0.284	0.294	-3.5	88	0.00
44 TM	Trichloroethene	0.389	0.417	-7.2	95	0.00
45 C	1,2-Dichloropropane	0.329	0.355	-7.9#	97	0.00
46 T	Dibromomethane	0.196	0.204	-4.1	91	0.00
47 T	Bromodichloromethane	0.500	0.521	-4.2	91	0.00
48 T	Methyl methacrylate	0.131	0.151	-15.3	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051321\
 Data File : VD069147.D
 Acq On : 13 May 2021 10:17
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 14 01:18:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	85	0.00
50 S	Toluene-d8	1.196	1.226	-2.5	98	0.00
51 T	4-Methyl-2-Pentanone	0.146	0.153	-4.8	87	0.00
52 CM	Toluene	0.899	0.983	-9.3#	93	0.00
53 T	t-1,3-Dichloropropene	0.439	0.453	-3.2	88	0.00
54 T	cis-1,3-Dichloropropene	0.525	0.537	-2.3	90	0.00
55 T	1,1,2-Trichloroethane	0.265	0.270	-1.9	90	0.00
56 T	Ethyl methacrylate	0.264	0.289	-9.5	88	0.00
57 T	1,3-Dichloropropane	0.428	0.451	-5.4	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.138	-5.3	88	0.00
59 T	2-Hexanone	0.097	0.112	-15.5	94	0.00
60 T	Dibromochloromethane	0.340	0.357	-5.0	92	0.00
61 T	1,2-Dibromoethane	0.262	0.267	-1.9	88	0.00
62 S	4-Bromofluorobenzene	0.385	0.387	-0.5	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.347	0.379	-9.2	97	0.00
65 PM	Chlorobenzene	1.036	1.100	-6.2	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.404	-3.3	92	0.00
67 C	Ethyl Benzene	1.783	1.985	-11.3#	95	0.00
68 T	m/p-Xylenes	0.699	0.790	-13.0	96	0.00
69 T	o-Xylene	0.622	0.694	-11.6	92	0.00
70 T	Styrene	1.085	1.220	-12.4	94	0.00
71 P	Bromoform	0.203	0.210	-3.4	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.495	3.884	-11.1	96	0.00
74 T	N-amyl acetate	0.572	0.596	-4.2	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.633	0.631	0.3	88	0.00
76 T	1,2,3-Trichloropropane	0.376	0.410	-9.0	90	0.00
77 T	Bromobenzene	0.829	0.899	-8.4	96	0.00
78 T	n-propylbenzene	4.217	4.759	-12.9	97	0.00
79 T	2-Chlorotoluene	2.346	2.545	-8.5	96	0.00
80 T	1,3,5-Trimethylbenzene	2.881	3.248	-12.7	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.186	0.187	-0.5	89	0.00
82 T	4-Chlorotoluene	2.499	2.688	-7.6	95	0.00
83 T	tert-Butylbenzene	2.401	2.757	-14.8	98	0.00
84 T	1,2,4-Trimethylbenzene	2.899	3.261	-12.5	96	0.00
85 T	sec-Butylbenzene	3.611	4.089	-13.2	97	0.00
86 T	p-Isopropyltoluene	3.172	3.685	-16.2	98	0.00
87 T	1,3-Dichlorobenzene	1.678	1.805	-7.6	96	0.00
88 T	1,4-Dichlorobenzene	1.662	1.755	-5.6	97	0.00
89 T	n-Butylbenzene	2.956	3.399	-15.0	96	0.00
90 T	Hexachloroethane	0.594	0.634	-6.7	96	0.00
91 T	1,2-Dichlorobenzene	1.441	1.517	-5.3	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.098	0.095	3.1	91	0.00
93 T	1,2,4-Trichlorobenzene	0.883	0.903	-2.3	90	0.00
94 T	Hexachlorobutadiene	0.530	0.579	-9.2	99	0.00
95 T	Naphthalene	1.480	1.519	-2.6	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051321\
Data File : VD069147.D
Acq On : 13 May 2021 10:17
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 14 01:18:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.756	0.786	-4.0	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6