

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD072995.D
 Acq On : 13 May 2022 17:53
 Operator : VA/SY
 Sample : N2816-01
 Misc : 6.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS002-1824-01

Quant Time: May 16 05:15:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	77471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	133564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	152653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	75466	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	67335	66.958	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.920%
35) Dibromofluoromethane	7.908	113	52954	56.452	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.900%
50) Toluene-d8	10.332	98	166533	49.492	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.620	95	71755	57.191	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD072995.D
Acq On : 13 May 2022 17:53
Operator : VA/SY
Sample : N2816-01
Misc : 6.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01

Quant Time: May 16 05:15:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

