

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD073002.D
 Acq On : 13 May 2022 21:15
 Operator : VA/SY
 Sample : N2816-12
 Misc : 4.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS004-0006-01

Quant Time: May 16 05:16:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	341583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	603623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	594300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	249952	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	187916	42.381	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.760%
35) Dibromofluoromethane	7.909	113	199778	47.125	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.260%
50) Toluene-d8	10.332	98	743774	48.911	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.820%
62) 4-Bromofluorobenzene	12.620	95	266951	47.080	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD073002.D
Acq On : 13 May 2022 21:15
Operator : VA/SY
Sample : N2816-12
Misc : 4.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS004-0006-01

Quant Time: May 16 05:16:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

