

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD073012.D
 Acq On : 14 May 2022 02:56
 Operator : VA/SY
 Sample : VD0513SBL02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0513SBL02

Quant Time: May 16 05:19:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	357767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	621935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	613743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	275607	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	188509	40.591	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.180%
35) Dibromofluoromethane	7.909	113	203612	46.615	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.240%
50) Toluene-d8	10.332	98	762081	48.639	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.280%
62) 4-Bromofluorobenzene	12.620	95	267991	45.871	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD073012.D
Acq On : 14 May 2022 02:56
Operator : VA/SY
Sample : VD0513SBL02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0513SBL02

Quant Time: May 16 05:19:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

