

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD073018.D
 Acq On : 14 May 2022 05:47
 Operator : VA/SY
 Sample : N2823-04
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SC-4

Quant Time: May 16 05:20:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	325278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	567682	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	560405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	265490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	181085	42.887	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.780%
35) Dibromofluoromethane	7.908	113	162640	40.794	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.580%
50) Toluene-d8	10.332	98	690326	48.270	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.540%
62) 4-Bromofluorobenzene	12.620	95	262360	49.199	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.400%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	27419	21.978	ug/l #	82
25) 2-Butanone	7.209	43	11384	9.253	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD073018.D
Acq On : 14 May 2022 05:47
Operator : VA/SY
Sample : N2823-04
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SC-4

Quant Time: May 16 05:20:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

