

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD073006.D
 Acq On : 13 May 2022 23:09
 Operator : VA/SY
 Sample : N2816-03MSD
 Misc : 5.26G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS002-1824-01MSD

Quant Time: May 16 05:17:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/16/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	311156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	510729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	472480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	228920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	177073	43.840	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	87.680%		
35) Dibromofluoromethane	7.908	113	186620	52.028	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	104.060%		
50) Toluene-d8	10.332	98	676735	52.596	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	105.200%		
62) 4-Bromofluorobenzene	12.620	95	245940	51.263	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	102.520%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.991	85	178557	41.527	ug/l	95
3) Chloromethane	2.209	50	169825	37.978	ug/l	96
4) Vinyl Chloride	2.344	62	168050	41.882	ug/l	98
5) Bromomethane	2.762	94	109360	44.128	ug/l	92
6) Chloroethane	2.914	64	103429	41.359	ug/l	97
7) Trichlorofluoromethane	3.267	101	324035	46.119	ug/l	98
8) Diethyl Ether	3.714	74	86060	42.724	ug/l	93
9) 1,1,2-Trichlorotrifluo...	4.091	101	192114	46.056	ug/l	97
10) Methyl Iodide	4.303	142	177539	42.410	ug/l	99
11) Tert butyl alcohol	5.220	59	71411	204.346	ug/l #	93
12) 1,1-Dichloroethene	4.067	96	174161	46.419	ug/l	86
13) Acrolein	3.914	56	11326	31.945	ug/l	99
14) Allyl chloride	4.708	41	280058	46.234	ug/l	92
15) Acrylonitrile	5.414	53	187668	196.039	ug/l	97
16) Acetone	4.150	43	174845	230.173	ug/l	94
17) Carbon Disulfide	4.409	76	478791	40.585	ug/l	100
18) Methyl Acetate	4.720	43	116791	53.154	ug/l	93
19) Methyl tert-butyl Ether	5.479	73	401192	44.604	ug/l	98
20) Methylene Chloride	4.961	84	225808	46.270	ug/l	87
21) trans-1,2-Dichloroethene	5.467	96	200129	46.509	ug/l	84
22) Diisopropyl ether	6.361	45	620845	46.592	ug/l	92
23) Vinyl Acetate	6.303	43	1131728m	153.746	ug/l	
24) 1,1-Dichloroethane	6.261	63	369754	46.186	ug/l	98
25) 2-Butanone	7.208	43	239635	203.624	ug/l	95
26) 2,2-Dichloropropane	7.202	77	290222	42.881	ug/l	97
27) cis-1,2-Dichloroethene	7.208	96	230001	46.669	ug/l	91
28) Bromochloromethane	7.538	49	143953	48.218	ug/l #	82
29) Tetrahydrofuran	7.555	42	147271	196.327	ug/l	92
30) Chloroform	7.702	83	385934	45.559	ug/l	99
31) Cyclohexane	7.979	56	302665	42.867	ug/l	90
32) 1,1,1-Trichloroethane	7.897	97	344956	47.780	ug/l	96
36) 1,1-Dichloropropene	8.108	75	287116	49.228	ug/l	98
37) Ethyl Acetate	7.291	43	95187	35.502	ug/l #	93
38) Carbon Tetrachloride	8.091	117	296683	49.003	ug/l	99
39) Methylcyclohexane	9.349	83	309624	45.767	ug/l	96
40) Benzene	8.344	78	818071	48.570	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
 Data File : VD073006.D
 Acq On : 13 May 2022 23:09
 Operator : VA/SY
 Sample : N2816-03MSD
 Misc : 5.26G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS002-1824-01MSD

Quant Time: May 16 05:17:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/16/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.526	41	63112	39.532	ug/l #	91
42) 1,2-Dichloroethane	8.420	62	224429	44.217	ug/l	97
43) Isopropyl Acetate	8.444	43	195984	39.603	ug/l #	92
44) Trichloroethene	9.108	130	215593	47.845	ug/l	93
45) 1,2-Dichloropropane	9.379	63	211393	47.791	ug/l	100
46) Dibromomethane	9.473	93	107443	44.034	ug/l	96
47) Bromodichloromethane	9.655	83	293959	46.832	ug/l	98
48) Methyl methacrylate	9.449	41	109813	45.359	ug/l	90
49) 1,4-Dioxane	9.449	88	23258	846.475	ug/l	98
51) 4-Methyl-2-Pentanone	10.220	43	545229	212.027	ug/l	95
52) Toluene	10.396	92	525083	49.976	ug/l	96
53) t-1,3-Dichloropropene	10.614	75	254303	44.108	ug/l	95
54) cis-1,3-Dichloropropene	10.085	75	310143	46.301	ug/l #	88
55) 1,1,2-Trichloroethane	10.791	97	148399	44.990	ug/l	99
56) Ethyl methacrylate	10.655	69	163040	41.981	ug/l #	88
57) 1,3-Dichloropropane	10.938	76	256419	45.300	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.932	63	451107	239.573	ug/l	97
59) 2-Hexanone	10.973	43	374863	216.643	ug/l	96
60) Dibromochloromethane	11.132	129	190703	45.927	ug/l	98
61) 1,2-Dibromoethane	11.238	107	139795	43.542	ug/l	98
64) Tetrachloroethene	10.867	164	199617	57.786	ug/l	97
65) Chlorobenzene	11.661	112	548835	50.222	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.732	131	210935	51.360	ug/l	97
67) Ethyl Benzene	11.738	91	988668	52.276	ug/l	99
68) m/p-Xylenes	11.843	106	772779	106.076	ug/l	98
69) o-Xylene	12.173	106	365554	53.728	ug/l	95
70) Styrene	12.185	104	624156	52.947	ug/l	97
71) Bromoform	12.349	173	102807	46.188	ug/l #	97
73) Isopropylbenzene	12.467	105	957504	53.760	ug/l	100
74) N-amyl acetate	12.279	43	169144	38.135	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.720	83	160657	44.685	ug/l	99
76) 1,2,3-Trichloropropane	12.767	75	133620m	52.346	ug/l	
77) Bromobenzene	12.749	156	211738	50.064	ug/l	93
78) n-propylbenzene	12.808	91	1175334	52.359	ug/l	99
79) 2-Chlorotoluene	12.896	91	669272	51.040	ug/l	99
80) 1,3,5-Trimethylbenzene	12.949	105	804763	53.288	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.514	75	47910	43.161	ug/l	92
82) 4-Chlorotoluene	12.990	91	702119	51.247	ug/l	100
83) tert-Butylbenzene	13.208	119	693208	53.990	ug/l	96
84) 1,2,4-Trimethylbenzene	13.255	105	788889	52.259	ug/l	99
85) sec-Butylbenzene	13.385	105	1007595	52.032	ug/l	99
86) p-Isopropyltoluene	13.502	119	834501	52.560	ug/l	99
87) 1,3-Dichlorobenzene	13.502	146	421139	49.429	ug/l	100
88) 1,4-Dichlorobenzene	13.579	146	415891	49.017	ug/l	99
89) n-Butylbenzene	13.826	91	778935	50.645	ug/l	99
90) Hexachloroethane	14.096	117	166275	51.121	ug/l	97
91) 1,2-Dichlorobenzene	13.873	146	365665	49.245	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.490	75	24243	42.537	ug/l	91
93) 1,2,4-Trichlorobenzene	15.143	180	205626	46.760	ug/l	100
94) Hexachlorobutadiene	15.249	225	98080	40.839	ug/l	98
95) Naphthalene	15.384	128	373613	46.210	ug/l	99
96) 1,2,3-Trichlorobenzene	15.573	180	176839	45.364	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD073006.D
Acq On : 13 May 2022 23:09
Operator : VA/SY
Sample : N2816-03MSD
Misc : 5.26G/5.00ml/MSVOA_D/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 16 05:17:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051322\
Data File : VD073006.D
Acq On : 13 May 2022 23:09
Operator : VA/SY
Sample : N2816-03MSD
Misc : 5.26G/5.00ml/MSVOA_D/SOIL
ALS Vial : 27 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022


Quant Time: May 16 05:17:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 10
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion:168 Resp: 311150
Ion Ratio Lower Upper
168 100
99 55.4 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#2
Dichlorodifluoromethane
Concen: 41.527 ug/l
RT: 1.991 min Scan# 12
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion: 85 Resp: 178557
Ion Ratio Lower Upper
85 100
87 33.8 15.7 47.0





#3
Chloromethane
Concen: 37.978 ug/l
RT: 2.209 min Scan# 49
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 50 Resp: 169825
Ion Ratio Lower Upper
50 100
52 31.7 27.4 41.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#4
Vinyl Chloride
Concen: 41.882 ug/l
RT: 2.344 min Scan# 72
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 62 Resp: 168050
Ion Ratio Lower Upper
62 100
64 32.7 25.4 38.0





#5
Bromomethane
Concen: 44.128 ug/l
RT: 2.762 min Scan# 144
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 94 Resp: 109360
Ion Ratio Lower Upper
94 100
96 88.9 77.6 116.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#6
Chloroethane
Concen: 41.359 ug/l
RT: 2.914 min Scan# 169
Delta R.T. -0.007 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 64 Resp: 103429
Ion Ratio Lower Upper
64 100
66 30.6 25.9 38.9





#7
Trichlorofluoromethane
Concen: 46.119 ug/l
RT: 3.267 min Scan# 21
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion:101 Resp: 324035
Ion Ratio Lower Upper
101 100
103 64.3 50.1 75.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#8
Diethyl Ether
Concen: 42.724 ug/l
RT: 3.714 min Scan# 305
Delta R.T. 0.005 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion: 74 Resp: 86060
Ion Ratio Lower Upper
74 100
45 98.5 52.6 157.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.056 ug/l
RT: 4.091 min Scan# 30
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion:101 Resp: 192114
Ion Ratio Lower Upper
101 100
85 44.0 36.4 54.6
151 73.9 57.3 85.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#10
Methyl Iodide
Concen: 42.410 ug/l
RT: 4.303 min Scan# 405
Delta R.T. 0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion:142 Resp: 177539
Ion Ratio Lower Upper
142 100
127 45.6 37.3 55.9
141 14.4 11.7 17.5





#11
Tert butyl alcohol
Concen: 204.346 ug/l
RT: 5.220 min Scan# 50
Delta R.T. 0.005 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 59 Resp: 7141
Ion Ratio Lower Upper
59 100
57 7.0 7.8 11.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#12
1,1-Dichloroethene
Concen: 46.419 ug/l
RT: 4.067 min Scan# 365
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 96 Resp: 174161
Ion Ratio Lower Upper
96 100
61 148.2 138.5 207.7
98 62.6 48.6 73.0





#13
Acrolein
Concen: 31.945 ug/l
RT: 3.914 min Scan# 31
Delta R.T. -0.007 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion: 56 Resp: 11320
Ion Ratio Lower Upper
56 100
55 72.2 57.1 85.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#14
Allyl chloride
Concen: 46.234 ug/l
RT: 4.708 min Scan# 474
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion: 41 Resp: 280058
Ion Ratio Lower Upper
41 100
39 66.6 59.3 88.9
76 35.6 25.7 38.5





#15
Acrylonitrile
Concen: 196.039 ug/l
RT: 5.414 min Scan# 595
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument : MSVOA_D
ClientSampleId : P007-SS002-1824-01MSD




Tgt Ion: 53 Resp: 187668
Ion Ratio Lower Upper
53 100
52 80.6 67.2 100.8
51 37.1 30.3 45.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#16
Acetone
Concen: 230.173 ug/l
RT: 4.150 min Scan# 379
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 43 Resp: 174845
Ion Ratio Lower Upper
43 100
58 30.7 22.1 33.1





#17
Carbon Disulfide
Concen: 40.585 ug/l
RT: 4.409 min Scan# 41
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 76 Resp: 47879
Ion Ratio Lower Upper
76 100
78 9.3 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#18
Methyl Acetate
Concen: 53.154 ug/l
RT: 4.720 min Scan# 476
Delta R.T. 0.005 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09



Tgt Ion: 43 Resp: 116791
Ion Ratio Lower Upper
43 100
74 24.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 44.604 ug/l

RT: 5.479 min Scan# 604

Delta R.T. 0.006 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 73 Resp: 401192

Ion Ratio Lower Upper

73 100





57 23.2 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#20

Methylene Chloride

Concen: 46.270 ug/l

RT: 4.961 min Scan# 517

Delta R.T. -0.001 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 84 Resp: 225808




Ion Ratio Lower Upper


84 100

49 117.9 112.8 169.2

51 37.9 32.6 48.8


86 62.3 47.9 71.9





#21
trans-1,2-Dichloroethene
Concen: 46.509 ug/l
RT: 5.467 min Scan# 604
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 96 Resp: 200129
Ion Ratio Lower Upper
96 100
61 123.9 117.9 176.9
98 55.2 49.4 74.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#22
Diisopropyl ether
Concen: 46.592 ug/l
RT: 6.361 min Scan# 755
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 45 Resp: 620845
Ion Ratio Lower Upper
45 100
43 51.8 46.6 70.0
87 27.5 18.4 27.6
59 11.3 8.6 13.0





#23
 Vinyl Acetate
 Concen: 153.746 ug/l m
 RT: 6.303 min Scan# 745
 Delta R.T. -0.000 min
 Lab File: VD073006.D
 Acq: 13 May 2022 23:09




Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS002-1824-01MSD



Tgt Ion: 43 Resp: 1131728
 Ion Ratio Lower Upper
 43 100
 86 10.9 7.1 10.7




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 05/16/2022
 Supervised By :Mahesh Dadoda 05/16/2022



#24
 1,1-Dichloroethane
 Concen: 46.186 ug/l
 RT: 6.261 min Scan# 738
 Delta R.T. 0.005 min
 Lab File: VD073006.D
 Acq: 13 May 2022 23:09

Tgt Ion: 63 Resp: 369754
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.6 10.9
 100 4.8 2.1 6.5









#25
2-Butanone
Concen: 203.624 ug/l
RT: 7.208 min Scan# 899
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion: 43 Resp: 23963
Ion Ratio Lower Upper
43 100
72 25.6 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#26
2,2-Dichloropropane
Concen: 42.881 ug/l
RT: 7.202 min Scan# 898
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion: 77 Resp: 290222
Ion Ratio Lower Upper
77 100
97 22.9 10.8 32.4









#27
cis-1,2-Dichloroethene
Concen: 46.669 ug/l
RT: 7.208 min Scan# 899
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion: 96 Resp: 23000
Ion Ratio Lower Upper
96 100
61 139.9 0.0 313.8
98 63.9 0.0 128.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#28
Bromochloromethane
Concen: 48.218 ug/l
RT: 7.538 min Scan# 955
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion: 49 Resp: 143953
Ion Ratio Lower Upper
49 100
129 1.1 0.0 4.0
130 80.5 52.6 79.0#





#29

Tetrahydrofuran

Concen: 196.327 ug/l

RT: 7.555 min Scan# 91

Delta R.T. -0.006 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 42 Resp: 14727





Ion	Ratio	Lower	Upper
42	100		
72	42.7	30.7	46.1
71	40.7	28.6	43.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#30

Chloroform

Concen: 45.559 ug/l

RT: 7.702 min Scan# 983




Delta R.T. -0.001 min


Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 83 Resp: 385934


Ion	Ratio	Lower	Upper
83	100		
85	65.3	52.7	79.1





#31
Cyclohexane
Concen: 42.867 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 56 Resp: 30266
Ion Ratio Lower Upper
56 100
69 34.4 25.4 38.2
84 89.8 63.0 94.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#32
1,1,1-Trichloroethane
Concen: 47.780 ug/l
RT: 7.897 min Scan# 1016
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 97 Resp: 344956
Ion Ratio Lower Upper
97 100
99 63.9 52.4 78.6
61 43.0 38.0 57.0





#33
1,2-Dichloroethane-d4
Concen: 43.840 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 65 Resp: 17707
Ion Ratio Lower Upper
65 100
67 53.3 0.0 101.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09



Tgt Ion:114 Resp: 510729
Ion Ratio Lower Upper
114 100
63 19.8 0.0 46.6
88 15.8 0.0 33.2





#35

Dibromofluoromethane

Concen: 52.028 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. -0.001 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 113 Resp: 186620





Ion	Ratio	Lower	Upper
113	100		
111	104.6	83.8	125.6
192	18.1	14.6	21.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#36

1,1-Dichloropropene

Concen: 49.228 ug/l

RT: 8.108 min Scan# 1052




Delta R.T. -0.001 min


Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 75 Resp: 287116


Ion	Ratio	Lower	Upper
75	100		
110	34.9	16.4	49.1
77	30.6	24.2	36.2





#37
Ethyl Acetate
Concen: 35.502 ug/l
RT: 7.291 min Scan# 913
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion: 43 Resp: 9518
Ion Ratio Lower Upper
43 100
61 15.3 9.8 14.8
70 10.7 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#38
Carbon Tetrachloride
Concen: 49.003 ug/l
RT: 8.091 min Scan# 1049
Delta R.T. -0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion:117 Resp: 296683
Ion Ratio Lower Upper
117 100
119 95.8 75.8 113.6
121 31.0 24.5 36.7





#39

Methylcyclohexane

Concen: 45.767 ug/l

RT: 9.349 min Scan# 11

Delta R.T. -0.001 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 83 Resp: 309624





Ion	Ratio	Lower	Upper
83	100		
55	81.8	63.2	94.8
98	51.9	38.2	57.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#40

Benzene

Concen: 48.570 ug/l

RT: 8.344 min Scan# 1092




Delta R.T. -0.006 min


Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 78 Resp: 818071


Ion	Ratio	Lower	Upper
78	100		
77	23.8	19.1	28.7





#41
Methacrylonitrile
Concen: 39.532 ug/l
RT: 7.526 min Scan# 911
Delta R.T. 0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 41 Resp: 6311
Ion Ratio Lower Upper
41 100
39 58.2 44.6 67.0
67 73.0 48.2 72.4
52 30.3 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#42
1,2-Dichloroethane
Concen: 44.217 ug/l
RT: 8.420 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 62 Resp: 224429
Ion Ratio Lower Upper
62 100
98 9.9 0.0 17.6





#43
Isopropyl Acetate
Concen: 39.603 ug/l
RT: 8.444 min Scan# 1109
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument : MSVOA_D
ClientSampleId : P007-SS002-1824-01MSD



Tgt Ion: 43 Resp: 195984
Ion Ratio Lower Upper
43 100
61 32.6 22.4 33.6
87 13.3 8.7 13.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#44
Trichloroethene
Concen: 47.845 ug/l
RT: 9.108 min Scan# 1222
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion:130 Resp: 215593
Ion Ratio Lower Upper
130 100
95 95.9 0.0 205.4





#45

1,2-Dichloropropane

Concen: 47.791 ug/l

RT: 9.379 min Scan# 1

Delta R.T. -0.000 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 63 Resp: 21139

Ion Ratio Lower Upper

63 100





65 32.8 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#46

Dibromomethane

Concen: 44.034 ug/l

RT: 9.473 min Scan# 1284

Delta R.T. 0.006 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09




Tgt Ion: 93 Resp: 107443


Ion Ratio Lower Upper

93 100

95 83.9 66.7 100.1

174 95.7 71.6 107.4





#47

Bromodichloromethane

Concen: 46.832 ug/l

RT: 9.655 min Scan# 1315

Delta R.T. -0.001 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion: 83 Resp: 293958





Ion	Ratio	Lower	Upper
83	100		
85	63.1	51.4	77.0
127	8.9	6.1	9.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#48

Methyl methacrylate

Concen: 45.359 ug/l

RT: 9.449 min Scan# 1280




Delta R.T. -0.001 min


Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 41 Resp: 109813

Ion	Ratio	Lower	Upper
41	100		
69	81.9	56.7	85.1
39	56.8	42.7	64.1









#49
1,4-Dioxane
Concen: 846.475 ug/l
RT: 9.449 min Scan# 1280
Delta R.T. -0.007 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion: 88 Resp: 2325
Ion Ratio Lower Upper
88 100
43 39.3 31.0 46.4
58 70.6 58.1 87.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#50
Toluene-d8
Concen: 52.596 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion: 98 Resp: 676735
Ion Ratio Lower Upper
98 100
100 66.9 52.4 78.6





#51
4-Methyl-2-Pentanone
Concen: 212.027 ug/l
RT: 10.220 min Scan# 1411
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 43 Resp: 545229
Ion Ratio Lower Upper
43 100
58 38.9 28.6 43.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#52
Toluene
Concen: 49.976 ug/l
RT: 10.396 min Scan# 1441
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 92 Resp: 525083
Ion Ratio Lower Upper
92 100
91 169.4 139.6 209.4





#53
t-1,3-Dichloropropene
Concen: 44.108 ug/l
RT: 10.614 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 75 Resp: 25430
Ion Ratio Lower Upper
75 100
77 33.3 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#54
cis-1,3-Dichloropropene
Concen: 46.301 ug/l
RT: 10.085 min Scan# 1388
Delta R.T. 0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 75 Resp: 310143
Ion Ratio Lower Upper
75 100
77 30.1 24.8 37.2
39 45.9 46.8 70.2#





#55
1,1,2-Trichloroethane
Concen: 44.990 ug/l
RT: 10.791 min Scan# 1508
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD






Tgt Ion: 97 Resp: 148399


Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.0	103.4
85	56.0	42.6	63.8
99	64.9	51.7	77.5

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022






#56
Ethyl methacrylate
Concen: 41.981 ug/l
RT: 10.655 min Scan# 1485
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09



Tgt Ion: 69 Resp: 163040

Ion	Ratio	Lower	Upper
69	100		
41	69.5	63.0	94.6
39	32.9	33.8	50.8





#57

1,3-Dichloropropane

Concen: 45.300 ug/l

RT: 10.938 min Scan# 1533

Delta R.T. -0.000 min

Lab File: VD073006.D


Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD



Tgt Ion: 76 Resp: 256419

Ion Ratio Lower Upper

76 100



78 32.9 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#58

2-Chloroethyl Vinyl ether

Concen: 239.573 ug/l

RT: 9.932 min Scan# 1362

Delta R.T. -0.000 min

Lab File: VD073006.D




Acq: 13 May 2022 23:09


Tgt Ion: 63 Resp: 451107

Ion Ratio Lower Upper

63 100


106 28.4 21.4 32.0





#59
2-Hexanone
Concen: 216.643 ug/l
RT: 10.973 min Scan# 1539
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument : MSVOA_D
ClientSampleId : P007-SS002-1824-01MSD



Tgt Ion: 43 Resp: 37486
Ion Ratio Lower Upper
43 100
58 53.7 25.3 75.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#60
Dibromochloromethane
Concen: 45.927 ug/l
RT: 11.132 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion:129 Resp: 190703
Ion Ratio Lower Upper
129 100
127 76.7 39.2 117.6





#61
1,2-Dibromoethane
Concen: 43.542 ug/l
RT: 11.238 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:107 Resp: 139795
Ion Ratio Lower Upper
107 100
109 95.1 74.2 111.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#62
4-Bromofluorobenzene
Concen: 51.263 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 95 Resp: 245940
Ion Ratio Lower Upper
95 100
174 73.5 0.0 152.8
176 71.9 0.0 145.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:117 Resp: 472480
Ion Ratio Lower Upper
117 100
82 53.0 45.3 67.9
119 33.0 26.6 40.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#64
Tetrachloroethene
Concen: 57.786 ug/l
RT: 10.867 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:164 Resp: 199617
Ion Ratio Lower Upper
164 100
166 125.5 98.5 147.7
129 96.0 74.6 111.8
131 93.8 72.0 108.0





#65
Chlorobenzene
Concen: 50.222 ug/l
RT: 11.661 min Scan# 1656
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:112 Resp: 548835
Ion Ratio Lower Upper
112 100
114 33.1 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 51.360 ug/l
RT: 11.732 min Scan# 1668
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:131 Resp: 210935
Ion Ratio Lower Upper
131 100
133 93.2 48.1 144.3
119 65.9 31.9 95.9





#67
Ethyl Benzene
Concen: 52.276 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 91 Resp: 988668
Ion Ratio Lower Upper
91 100
106 30.8 24.2 36.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#68
m/p-Xylenes
Concen: 106.076 ug/l
RT: 11.843 min Scan# 1687
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09



Tgt Ion:106 Resp: 772779
Ion Ratio Lower Upper
106 100
91 201.2 163.7 245.5









#69
o-Xylene
Concen: 53.728 ug/l
RT: 12.173 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion:106 Resp: 365554
Ion Ratio Lower Upper
106 100
91 212.2 110.4 331.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#70
Styrene
Concen: 52.947 ug/l
RT: 12.185 min Scan# 1745
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion:104 Resp: 624156
Ion Ratio Lower Upper
104 100
78 49.4 41.8 62.8
103 53.9 44.2 66.4





#71
Bromoform
Concen: 46.188 ug/l
RT: 12.349 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:173 Resp: 102807
Ion Ratio Lower Upper
173 100
175 49.8 23.9 71.8
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:152 Resp: 228920
Ion Ratio Lower Upper
152 100
115 59.6 31.3 93.8
150 173.5 0.0 348.6





#73
Isopropylbenzene
Concen: 53.760 ug/l
RT: 12.467 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion:105 Resp: 957504
Ion Ratio Lower Upper
105 100
120 26.0 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#74
N-ethyl acetate
Concen: 38.135 ug/l
RT: 12.279 min Scan# 1761
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion: 43 Resp: 169144
Ion Ratio Lower Upper
43 100
70 42.0 31.3 46.9
55 25.2 20.0 30.0
61 24.3 17.8 26.8





#75
1,1,2,2-Tetrachloroethane
Concen: 44.685 ug/l
RT: 12.720 min Scan# 1836
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 83 Resp: 16065
Ion Ratio Lower Upper
83 100
131 10.9 5.0 14.9
85 65.6 32.6 98.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#76
1,2,3-Trichloropropane
Concen: 52.346 ug/l m
RT: 12.767 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion: 75 Resp: 133620
Ion Ratio Lower Upper
75 100
77 303.6 0.0 0.0#





#77
Bromobenzene
Concen: 50.064 ug/l
RT: 12.749 min Scan# 1841
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion: 156 Resp: 211738
Ion Ratio Lower Upper
156 100
77 191.6 102.8 308.3
158 95.6 48.7 146.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#78
n-propylbenzene
Concen: 52.359 ug/l
RT: 12.808 min Scan# 1851
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09



Tgt Ion: 91 Resp: 1175334
Ion Ratio Lower Upper
91 100
120 22.3 10.9 32.9









#79
2-Chlorotoluene
Concen: 51.040 ug/l
RT: 12.896 min Scan# 1866
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD

Tgt Ion: 91 Resp: 669272
Ion Ratio Lower Upper
91 100
126 33.2 16.2 48.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#80
1,3,5-Trimethylbenzene
Concen: 53.288 ug/l
RT: 12.949 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion:105 Resp: 804763
Ion Ratio Lower Upper
105 100
120 48.8 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 43.161 ug/l
RT: 12.514 min Scan# 1801
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion: 75 Resp: 47910
Ion Ratio Lower Upper
75 100
53 98.5 86.5 129.7
89 45.9 33.9 50.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#82
4-Chlorotoluene
Concen: 51.247 ug/l
RT: 12.990 min Scan# 1882
Delta R.T. -0.001 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09


Tgt Ion: 91 Resp: 702119
Ion Ratio Lower Upper
91 100
126 32.9 16.5 49.5





#83
tert-Butylbenzene
Concen: 53.990 ug/l
RT: 13.208 min Scan# 1919
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:119 Resp: 693208
Ion Ratio Lower Upper
119 100
91 65.0 34.8 104.5
134 26.2 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#84
1,2,4-Trimethylbenzene
Concen: 52.259 ug/l
RT: 13.255 min Scan# 1927
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:105 Resp: 788889
Ion Ratio Lower Upper
105 100
120 45.0 22.3 66.8





#85
 sec-Butylbenzene
 Concen: 52.032 ug/l
 RT: 13.385 min Scan# 1949
 Delta R.T. -0.000 min
 Lab File: VD073006.D
 Acq: 13 May 2022 23:09




Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS002-1824-01MSD




Tgt Ion:105 Resp: 100759
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.8 29.3

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 05/16/2022
 Supervised By :Mahesh Dadoda 05/16/2022




#86
 p-Isopropyltoluene
 Concen: 52.560 ug/l
 RT: 13.502 min Scan# 1969
 Delta R.T. 0.000 min
 Lab File: VD073006.D
 Acq: 13 May 2022 23:09




Tgt Ion:119 Resp: 834501
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.6
 91 23.7 12.1 36.3





#87
1,3-Dichlorobenzene
Concen: 49.429 ug/l
RT: 13.502 min Scan# 1969
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument : MSVOA_D
ClientSampleId : P007-SS002-1824-01MSD




Tgt Ion:146 Resp: 421139
Ion Ratio Lower Upper
146 100
111 41.0 20.4 61.3
148 63.8 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#88
1,4-Dichlorobenzene
Concen: 49.017 ug/l
RT: 13.579 min Scan# 1982
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:146 Resp: 415891
Ion Ratio Lower Upper
146 100
111 40.1 20.3 60.9
148 64.3 31.7 95.1





#89
n-Butylbenzene
Concen: 50.645 ug/l
RT: 13.826 min Scan# 2024
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion: 91 Resp: 778935
Ion Ratio Lower Upper
91 100
92 53.7 26.6 79.8
134 25.1 12.3 36.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#90
Hexachloroethane
Concen: 51.121 ug/l
RT: 14.096 min Scan# 2070
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion:117 Resp: 166275
Ion Ratio Lower Upper
117 100
201 64.2 33.4 100.2





#91

1,2-Dichlorobenzene

Concen: 49.245 ug/l

RT: 13.873 min Scan# 2032

Delta R.T. -0.000 min

Lab File: VD073006.D

Acq: 13 May 2022 23:09

Instrument :

MSVOA_D

ClientSampleId :

P007-SS002-1824-01MSD

Tgt Ion:146 Resp: 36566





Ion	Ratio	Lower	Upper
146	100		
111	42.0	21.1	63.3
148	63.7	31.7	95.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2022

Supervised By :Mahesh Dadoda 05/16/2022



#92

1,2-Dibromo-3-Chloropropane

Concen: 42.537 ug/l

RT: 14.490 min Scan# 2137




Delta R.T. -0.001 min


Lab File: VD073006.D

Acq: 13 May 2022 23:09

Tgt Ion: 75 Resp: 24243


Ion	Ratio	Lower	Upper
75	100		
155	83.8	36.8	110.3
157	103.7	48.4	145.0





#93
1,2,4-Trichlorobenzene
Concen: 46.760 ug/l
RT: 15.143 min Scan# 21
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD




Tgt Ion:180 Resp: 205620
Ion Ratio Lower Upper
180 100
182 94.9 47.3 142.0
145 31.5 16.2 48.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022




#94
Hexachlorobutadiene
Concen: 40.839 ug/l
RT: 15.249 min Scan# 2266
Delta R.T. 0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Tgt Ion:225 Resp: 98080
Ion Ratio Lower Upper
225 100
223 63.4 31.1 93.2
227 62.2 32.4 97.0





#95
Naphthalene
Concen: 46.210 ug/l
RT: 15.384 min Scan# 21
Delta R.T. 0.005 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09




Instrument :
MSVOA_D
ClientSampleId :
P007-SS002-1824-01MSD



Tgt Ion:128 Resp: 37361
Ion Ratio Lower Upper
128 100
127 13.3 10.3 15.5
129 10.7 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/16/2022
Supervised By :Mahesh Dadoda 05/16/2022



#96
1,2,3-Trichlorobenzene
Concen: 45.364 ug/l
RT: 15.573 min Scan# 2321
Delta R.T. -0.000 min
Lab File: VD073006.D
Acq: 13 May 2022 23:09

Tgt Ion:180 Resp: 176839
Ion Ratio Lower Upper
180 100
182 92.9 47.6 142.9
145 33.4 17.1 51.3

