

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051324\
 Data File : VD078911.D
 Acq On : 13 May 2024 13:38
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: May 14 06:29:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 14 06:21:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	123671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	285862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	276712	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	124447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	67468	58.160	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.320%
35) Dibromofluoromethane	7.805	113	95008	49.928	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.860%
50) Toluene-d8	10.269	98	359032	50.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.660%
62) 4-Bromofluorobenzene	12.569	95	96665	48.236	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.480%
Target Compounds						
20) Methylene Chloride	4.799	84	15981	5.682	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051324\
Data File : VD078911.D
Acq On : 13 May 2024 13:38
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: May 14 06:29:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
Quant Title : SW846 8260
QLast Update : Tue May 14 06:21:28 2024
Response via : Initial Calibration

