

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD051419\
 Data File : VD062367.D
 Acq On : 15 May 2019 2:09
 Operator : FY/SY
 Sample : K2756-16RE
 Misc : 5.11g/5mL/MSVOA_D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P021-SS008-0002-01RE

Quant Time: May 15 03:47:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	71311	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.23	114	124835	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.37	117	97899	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	41150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	74762	112.81	ug/l	0.04
Spiked Amount 50.000			Recovery	=	225.62%	
35) Dibromofluoromethane	6.94	113	57376	55.39	ug/l	0.05
Spiked Amount 50.000			Recovery	=	110.78%	
50) Toluene-d8	10.42	98	70316	28.16	ug/l	0.02
Spiked Amount 50.000			Recovery	=	56.32%	
62) 4-Bromofluorobenzene	13.56	95	32617	32.52	ug/l	0.02
Spiked Amount 50.000			Recovery	=	65.04%	
Target Compounds						
16) Acetone	3.23	43	25126	180.794	ug/l #	Qvalue 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD051419\
Data File : VD062367.D
Acq On : 15 May 2019 2:09
Operator : FY/SY
Sample : K2756-16RE
Misc : 5.11g/5mL/MSVOA_D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P021-SS008-0002-01RE

Quant Time: May 15 03:47:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.06 min Scan# 573

Delta R.T. 0.04 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09

Instrument :

MSVOA_D

ClientSampleId :


P021-SS008-0002-01RE

Tgt Ion:168 Resp: 71311

Ion Ratio Lower Upper




168 100

99 72.0 46.5 69.7#



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#16

Acetone

Concen: 180.794 ug/l

RT: 3.23 min Scan# 184

Delta R.T. 0.04 min

Lab File: VD062367.D


Acq: 15 May 2019 2:09

Tgt Ion: 43 Resp: 25126

Ion Ratio Lower Upper



43 100


58 14.6 19.9 29.9#



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC





#33

1,2-Dichloroethane-d4

Concen: 112.814 ug/l

RT: 7.46 min Scan# 614

Delta R.T. 0.04 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09

Instrument :

MSVOA_D

ClientSampleId :


P021-SS008-0002-01RE

Tgt Ion: 65 Resp: 74762

Ion Ratio Lower Upper




65 100

67 45.4 0.0 98.0



Abundance Ion 64.95 (64.65 to 65.65): VDC

Ion 66.90 (66.60 to 67.60): VDC



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.23 min Scan# 692

Delta R.T. 0.04 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09


Tgt Ion: 114 Resp: 124835

Ion Ratio Lower Upper

114 100

63 22.3 0.0 32.4



88 17.3 0.0 33.6




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 62.95 (62.65 to 63.65): VDC

Ion 87.90 (87.60 to 88.60): VDC





#35

Dibromofluoromethane

Concen: 55.395 ug/l

RT: 6.94 min Scan# 561

Delta R.T. 0.05 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09

Instrument :

MSVOA_D

ClientSampleId :

P021-SS008-0002-01RE


Tgt Ion:113 Resp: 57376

Ion Ratio Lower Upper

113 100

111 94.1 78.5 117.7




192 14.3 13.0 19.4



Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V



#50

Toluene-d8

Concen: 28.157 ug/l

RT: 10.42 min Scan# 914

Delta R.T. 0.02 min

Lab File: VD062367.D


Acq: 15 May 2019 2:09

Tgt Ion: 98 Resp: 70316

Ion Ratio Lower Upper



98 100


100 65.0 53.8 80.8



Abundance Ion 98.05 (97.75 to 98.75): V

Ion 100.05 (99.75 to 100.75): V





#62

4-Bromofluorobenzene

Concen: 32.525 ug/l

RT: 13.56 min Scan# 1234

Delta R.T. 0.02 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09

Instrument :

MSVOA_D

ClientSampleId :

P021-SS008-0002-01RE


Tgt Ion: 95 Resp: 32617

Ion Ratio Lower Upper

95 100

174 99.7 0.0 181.8




176 91.1 0.0 174.4



Abundance Ion 95.00 (94.70 to 95.70): V

Ion 173.90 (173.60 to 174.60): V

Ion 175.90 (175.60 to 176.60): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 12.37 min Scan# 1113

Delta R.T. 0.02 min

Lab File: VD062367.D

Acq: 15 May 2019 2:09


Tgt Ion: 117 Resp: 97899

Ion Ratio Lower Upper

117 100

82 58.1 36.2 54.2#



119 29.7 26.2 39.4




Abundance Ion 116.95 (116.65 to 117.65): V

Ion 82.05 (81.75 to 82.75): V

Ion 118.95 (118.65 to 119.65): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 14.52 min Scan# 1331

Delta R.T. 0.01 min

Lab File: VD062367.D


Acq: 15 May 2019 2:09

Instrument :

MSVOA_D

ClientSampleId :

P021-SS008-0002-01RE



Tgt Ion:152 Resp: 41150

Ion Ratio Lower Upper

152 100

115 68.5 30.9 92.7

150 157.5 0.0 314.8

