

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD051419\
 Data File : VD062374.D
 Acq On : 15 May 2019 5:23
 Operator : FY/SY
 Sample : K2735-17RE
 Misc : 5.05g/5mL/MSVOA_D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P026-SS017-1824-01RE

Quant Time: May 15 06:40:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 14:26:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	33044	50.00	ug/l	0.05
34) 1,4-Difluorobenzene	8.24	114	58798	50.00	ug/l	0.05
63) Chlorobenzene-d5	12.38	117	45828	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.53	152	16840	50.00	ug/l	0.02

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	31748	103.39	ug/l	0.05
Spiked Amount	50.000		Recovery	=	206.78%	
35) Dibromofluoromethane	6.94	113	24271	49.75	ug/l	0.05
Spiked Amount	50.000		Recovery	=	99.50%	
50) Toluene-d8	10.43	98	33877	28.80	ug/l	0.03
Spiked Amount	50.000		Recovery	=	57.60%	
62) 4-Bromofluorobenzene	13.57	95	17637	37.34	ug/l	0.02
Spiked Amount	50.000		Recovery	=	74.68%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD051419\
Data File : VD062374.D
Acq On : 15 May 2019 5:23
Operator : FY/SY
Sample : K2735-17RE
Misc : 5.05g/5mL/MSVOA_D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P026-SS017-1824-01RE

Quant Time: May 15 06:40:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D050719S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 14:26:29 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.07 min Scan# 574

Delta R.T. 0.05 min

Lab File: VD062374.D

Acq: 15 May 2019 5:23

Instrument :

MSVOA_D

ClientSampleId :


P026-SS017-1824-01RE

Tgt Ion:168 Resp: 33044

Ion Ratio Lower Upper




168 100

99 62.8 46.5 69.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#33

1,2-Dichloroethane-d4

Concen: 103.386 ug/l

RT: 7.47 min Scan# 615

Delta R.T. 0.05 min

Lab File: VD062374.D


Acq: 15 May 2019 5:23

Tgt Ion: 65 Resp: 31748

Ion Ratio Lower Upper



65 100


67 48.2 0.0 98.0



Abundance Ion 64.95 (64.65 to 65.65): VDC

Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.24 min Scan# 693

Delta R.T. 0.05 min

Lab File: VD062374.D

Acq: 15 May 2019 5:23

Instrument :

MSVOA_D

ClientSampleId :

P026-SS017-1824-01RE



Tgt Ion:114 Resp: 58798

Ion Ratio Lower Upper

114 100

63 18.8 0.0 32.4



88 17.8 0.0 33.6



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 62.95 (62.65 to 63.65): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 49.751 ug/l

RT: 6.94 min Scan# 561

Delta R.T. 0.05 min

Lab File: VD062374.D

Acq: 15 May 2019 5:23



Tgt Ion:113 Resp: 24271

Ion Ratio Lower Upper

113 100

111 104.5 78.5 117.7


192 20.7 13.0 19.4#




Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V





#50

Toluene-d8

Concen: 28.801 ug/l

RT: 10.43 min Scan# 915

Delta R.T. 0.03 min

Lab File: VD062374.D

Acq: 15 May 2019 5:23

Instrument :

MSVOA_D

ClientSampleId :


P026-SS017-1824-01RE

Tgt Ion: 98 Resp: 33877

Ion Ratio Lower Upper

98 100

100 66.4 53.8 80.8



Abundance Ion 98.05 (97.75 to 98.75): VD062139.D (-905) (-)

Ion 100.05 (99.75 to 100.75): VD062139.D (-905) (-)

10.43

15000

10000



5000

0

Time-->

10.40

10.50



#62

4-Bromofluorobenzene

Concen: 37.340 ug/l

RT: 13.57 min Scan# 1234

Delta R.T. 0.02 min

Lab File: VD062374.D

Acq: 15 May 2019 5:23


Tgt Ion: 95 Resp: 17637

Ion Ratio Lower Upper

95 100

174 85.5 0.0 181.8

176 87.2 0.0 174.4



Abundance Ion 95.00 (94.70 to 95.70): VD062139.D (-1227) (-)

Ion 173.90 (173.60 to 174.60): VD062139.D (-1227) (-)

Ion 175.90 (175.60 to 176.60): VD062139.D (-1227) (-)

13.57

15000

10000

5000


0


Time-->

13.50

13.55

13.60









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 12.38 min Scan# 1113
Delta R.T. 0.02 min
Lab File: VD062374.D
Acq: 15 May 2019 5:23

Instrument :
MSVOA_D
ClientSampleId :
P026-SS017-1824-01RE

Tgt Ion:117 Resp: 45828
Ion Ratio Lower Upper
117 100
82 59.4 36.2 54.2#
119 31.8 26.2 39.4




Abundance Ion 116.95 (116.65 to 117.65): V
Ion 82.05 (81.75 to 82.75): V
Ion 118.95 (118.65 to 119.65): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 14.53 min Scan# 1332
Delta R.T. 0.02 min
Lab File: VD062374.D
Acq: 15 May 2019 5:23

Tgt Ion:152 Resp: 16840
Ion Ratio Lower Upper
152 100
115 48.5 30.9 92.7
150 149.8 0.0 314.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 115.00 (114.70 to 115.70): V
Ion 149.90 (149.60 to 150.60): V

