

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051421\
 Data File : VD069176.D
 Acq On : 14 May 2021 14:58
 Operator : VA/SY
 Sample : M2387-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 41789

Quant Time: May 15 02:34:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	6105	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	9367	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.644	117	10433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	5163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	5376	87.84	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 175.68%#	
35) Dibromofluoromethane	7.914	113	3312	54.33	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.66%	
50) Toluene-d8	10.338	98	11262	50.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.56%	
62) 4-Bromofluorobenzene	12.626	95	6325	87.76	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 175.52%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051421\
Data File : VD069176.D
Acq On : 14 May 2021 14:58
Operator : VA/SY
Sample : M2387-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
41789

Quant Time: May 15 02:34:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

