

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051421\
 Data File : VD069186.D
 Acq On : 14 May 2021 19:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 15 02:37:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.532	0.460	13.5	86	0.00
3 P	Chloromethane	0.611	0.628	-2.8	106	0.00
4 C	Vinyl Chloride	0.805	0.764	5.1#	91	0.00
5 T	Bromomethane	0.606	0.567	6.4	99	0.00
6 T	Chloroethane	0.518	0.549	-6.0	101	0.00
7 T	Trichlorofluoromethane	0.982	0.901	8.2	92	0.00
8 T	Diethyl Ether	0.224	0.258	-15.2	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.524	5.9	92	0.00
10 T	Methyl Iodide	0.531	0.515	3.0	82	0.00
11 T	Tert butyl alcohol	0.034	0.031	8.8	112	0.00
12 CM	1,1-Dichloroethene	0.500	0.498	0.4#	93	0.00
13 T	Acrolein	0.041	0.044	-7.3	99	0.00
14 T	Allyl chloride	0.549	0.561	-2.2	96	0.00
15 T	Acrylonitrile	0.084	0.095	-13.1	105	0.00
16 T	Acetone	0.077	0.072	6.5	85	0.00
17 T	Carbon Disulfide	1.658	1.585	4.4	93	0.00
18 T	Methyl Acetate	0.187	0.211	-12.8	104	0.00
19 T	Methyl tert-butyl Ether	0.908	1.067	-17.5	105	0.00
20 T	Methylene Chloride	0.813	0.784	3.6	113	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.616	-3.9	98	0.00
22 T	Diisopropyl ether	1.096	1.253	-14.3	103	0.00
23 T	Vinyl Acetate	0.647	0.775	-19.8	103	0.00
24 P	1,1-Dichloroethane	0.947	0.987	-4.2	100	0.00
25 T	2-Butanone	0.104	0.113	-8.7	100	0.00
26 T	2,2-Dichloropropane	0.885	0.847	4.3	91	-0.01
27 T	cis-1,2-Dichloroethene	0.644	0.691	-7.3	103	0.00
28 T	Bromochloromethane	0.338	0.357	-5.6	102	0.00
29 T	Tetrahydrofuran	0.057	0.067	-17.5	104	0.00
30 C	Chloroform	1.066	1.124	-5.4#	101	0.00
31 T	Cyclohexane	0.743	0.685	7.8	93	0.00
32 T	1,1,1-Trichloroethane	0.967	0.975	-0.8	97	0.00
33 S	1,2-Dichloroethane-d4	0.501	0.496	1.0	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.325	0.309	4.9	101	0.00
36 T	1,1-Dichloropropene	0.486	0.479	1.4	96	0.00
37 T	Ethyl Acetate	0.156	0.170	-9.0	106	0.00
38 T	Carbon Tetrachloride	0.543	0.506	6.8	93	0.00
39 T	Methylcyclohexane	0.513	0.487	5.1	90	0.00
40 TM	Benzene	1.392	1.415	-1.7	101	0.00
41 T	Methacrylonitrile	0.083	0.090	-8.4	109	0.00
42 TM	1,2-Dichloroethane	0.380	0.403	-6.1	102	0.00
43 T	Isopropyl Acetate	0.284	0.319	-12.3	106	0.00
44 TM	Trichloroethene	0.389	0.390	-0.3	98	0.00
45 C	1,2-Dichloropropane	0.329	0.341	-3.6#	103	0.00
46 T	Dibromomethane	0.196	0.206	-5.1	102	0.00
47 T	Bromodichloromethane	0.500	0.516	-3.2	99	0.00
48 T	Methyl methacrylate	0.131	0.144	-9.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051421\
 Data File : VD069186.D
 Acq On : 14 May 2021 19:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 15 02:37:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	101	0.00
50 S	Toluene-d8	1.196	1.130	5.5	100	0.00
51 T	4-Methyl-2-Pentanone	0.146	0.167	-14.4	105	0.00
52 CM	Toluene	0.899	0.947	-5.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.439	0.469	-6.8	101	0.00
54 T	cis-1,3-Dichloropropene	0.525	0.549	-4.6	102	0.00
55 T	1,1,2-Trichloroethane	0.265	0.277	-4.5	102	0.00
56 T	Ethyl methacrylate	0.264	0.309	-17.0	104	0.00
57 T	1,3-Dichloropropane	0.428	0.461	-7.7	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.155	-18.3	109	0.00
59 T	2-Hexanone	0.097	0.112	-15.5	104	0.00
60 T	Dibromochloromethane	0.340	0.357	-5.0	102	0.00
61 T	1,2-Dibromoethane	0.262	0.275	-5.0	100	0.00
62 S	4-Bromofluorobenzene	0.385	0.374	2.9	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	106	0.00
64 T	Tetrachloroethene	0.347	0.333	4.0	96	0.00
65 PM	Chlorobenzene	1.036	1.040	-0.4	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.393	-0.5	101	0.00
67 C	Ethyl Benzene	1.783	1.805	-1.2#	97	0.00
68 T	m/p-Xylenes	0.699	0.711	-1.7	98	0.00
69 T	o-Xylene	0.622	0.658	-5.8	98	0.00
70 T	Styrene	1.085	1.157	-6.6	101	0.00
71 P	Bromoform	0.203	0.205	-1.0	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.495	3.530	-1.0	94	0.00
74 T	N-amyl acetate	0.572	0.642	-12.2	104	0.00
75 P	1,1,2,2-Tetrachloroethane	0.633	0.661	-4.4	99	0.00
76 T	1,2,3-Trichloropropane	0.376	0.402	-6.9	95	0.00
77 T	Bromobenzene	0.829	0.875	-5.5	101	0.00
78 T	n-propylbenzene	4.217	4.273	-1.3	94	0.00
79 T	2-Chlorotoluene	2.346	2.414	-2.9	98	0.00
80 T	1,3,5-Trimethylbenzene	2.881	2.973	-3.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.186	0.190	-2.2	98	0.00
82 T	4-Chlorotoluene	2.499	2.514	-0.6	96	0.00
83 T	tert-Butylbenzene	2.401	2.527	-5.2	97	0.00
84 T	1,2,4-Trimethylbenzene	2.899	3.049	-5.2	96	0.00
85 T	sec-Butylbenzene	3.611	3.684	-2.0	94	0.00
86 T	p-Isopropyltoluene	3.172	3.220	-1.5	92	0.00
87 T	1,3-Dichlorobenzene	1.678	1.684	-0.4	97	0.00
88 T	1,4-Dichlorobenzene	1.662	1.659	0.2	98	0.00
89 T	n-Butylbenzene	2.956	2.949	0.2	90	0.00
90 T	Hexachloroethane	0.594	0.569	4.2	93	0.00
91 T	1,2-Dichlorobenzene	1.441	1.476	-2.4	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.098	0.096	2.0	99	0.00
93 T	1,2,4-Trichlorobenzene	0.883	0.868	1.7	94	0.00
94 T	Hexachlorobutadiene	0.530	0.497	6.2	91	0.00
95 T	Naphthalene	1.480	1.669	-12.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051421\
Data File : VD069186.D
Acq On : 14 May 2021 19:41
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 15 02:37:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.756	0.778	-2.9	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6