

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD051520\
 Data File : VD065750.D
 Acq On : 15 May 2020 14:08
 Operator : VA/SY
 Sample : L2503-06
 Misc : 7.70G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-051420

Quant Time: May 16 04:08:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D051320S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 13 14:28:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	217637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	348625	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	283316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	97108	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.34	65	109234	62.28	ug/l	0.00
Spiked Amount			Recovery	=	124.56%	
35) Dibromofluoromethane	7.92	113	110269	55.83	ug/l	0.00
Spiked Amount			Recovery	=	111.66%	
50) Toluene-d8	10.34	98	372625	49.56	ug/l	0.00
Spiked Amount			Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.64	95	110240	43.29	ug/l	0.00
Spiked Amount			Recovery	=	86.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.16	43	18385	51.106	ug/l	92
39) Methylcyclohexane	9.36	83	15889	4.095	ug/l	96
52) Toluene	10.41	92	78740	14.437	ug/l	99
67) Ethyl Benzene	11.75	91	21343	2.242	ug/l	97
68) m/p-Xylenes	11.85	106	40255	11.072	ug/l	99
69) o-Xylene	12.18	106	13216	3.905	ug/l	92
80) 1,3,5-Trimethylbenzene	12.96	105	5888	1.096	ug/l	92
84) 1,2,4-Trimethylbenzene	13.28	105	11670	2.203	ug/l	96
95) Naphthalene	15.41	128	61302	19.005	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD051520\
Data File : VD065750.D
Acq On : 15 May 2020 14:08
Operator : VA/SY
Sample : L2503-06
Misc : 7.70G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-051420

Quant Time: May 16 04:08:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D051320S.M
Quant Title : SW846 8260
QLast Update : Wed May 13 14:28:42 2020
Response via : Initial Calibration

