

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051622\
 Data File : VD073039.D
 Acq On : 16 May 2022 17:08
 Operator : VA/SY
 Sample : N2816-19
 Misc : 5.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS011-0006-01

Quant Time: May 17 03:29:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	317584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	560732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	538073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	224111	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	166898	40.485	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.960%
35) Dibromofluoromethane	7.908	113	185418	47.083	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.160%
50) Toluene-d8	10.332	98	674322	47.735	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.620	95	226199	42.944	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051622\
Data File : VD073039.D
Acq On : 16 May 2022 17:08
Operator : VA/SY
Sample : N2816-19
Misc : 5.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS011-0006-01

Quant Time: May 17 03:29:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

