

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051622\
 Data File : VD073031.D
 Acq On : 16 May 2022 12:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00 G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 17 03:27:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.691	0.665	3.8	95	0.00
3 P	Chloromethane	0.719	0.607	15.6	86	0.00
4 C	Vinyl Chloride	0.645	0.622	3.6#	92	0.00
5 T	Bromomethane	0.433	0.402	7.2	98	-0.01
6 T	Chloroethane	0.402	0.374	7.0	94	0.00
7 T	Trichlorofluoromethane	1.129	1.189	-5.3	104	0.00
8 T	Diethyl Ether	0.324	0.320	1.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.670	0.731	-9.1	108	0.00
10 T	Methyl Iodide	0.673	0.687	-2.1	95	0.00
11 T	Tert butyl alcohol	0.056	0.056	0.0	113	0.00
12 CM	1,1-Dichloroethene	0.603	0.635	-5.3#	104	0.00
13 T	Acrolein	0.045	0.013	71.1#	34#	0.00
14 T	Allyl chloride	0.973	0.992	-2.0	99	0.00
15 T	Acrylonitrile	0.154	0.152	1.3	97	0.00
16 T	Acetone	0.136	0.131	3.7	92	0.00
17 T	Carbon Disulfide	1.896	1.771	6.6	92	0.00
18 T	Methyl Acetate	0.383	0.345	9.9	97	0.00
19 T	Methyl tert-butyl Ether	1.445	1.476	-2.1	97	0.00
20 T	Methylene Chloride	1.004	0.749	25.4#	95	0.00
21 T	trans-1,2-Dichloroethene	0.691	0.710	-2.7	99	0.00
22 T	Diisopropyl ether	2.141	2.129	0.6	94	0.00
23 T	Vinyl Acetate	1.183	1.196	-1.1	94	0.00
24 P	1,1-Dichloroethane	1.286	1.303	-1.3	101	0.00
25 T	2-Butanone	0.189	0.191	-1.1	96	0.00
26 T	2,2-Dichloropropane	1.088	1.173	-7.8	105	0.00
27 T	cis-1,2-Dichloroethene	0.792	0.797	-0.6	99	0.00
28 T	Bromochloromethane	0.480	0.455	5.2	89	0.00
29 T	Tetrahydrofuran	0.121	0.120	0.8	96	0.00
30 C	Chloroform	1.361	1.361	0.0	100	0.00
31 T	Cyclohexane	1.135	1.146	-1.0	98	0.00
32 T	1,1,1-Trichloroethane	1.160	1.253	-8.0	106	0.00
33 S	1,2-Dichloroethane-d4	0.649	0.557	14.2	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.351	0.351	0.0	96	0.00
36 T	1,1-Dichloropropene	0.571	0.648	-13.5	103	0.00
37 T	Ethyl Acetate	0.262	0.275	-5.0	95	0.00
38 T	Carbon Tetrachloride	0.593	0.678	-14.3	106	0.00
39 T	Methylcyclohexane	0.662	0.766	-15.7	105	0.00
40 TM	Benzene	1.649	1.775	-7.6	100	0.00
41 T	Methacrylonitrile	0.156	0.160	-2.6	87	0.00
42 TM	1,2-Dichloroethane	0.497	0.494	0.6	93	0.00
43 T	Isopropyl Acetate	0.484	0.506	-4.5	95	0.00
44 TM	Trichloroethene	0.441	0.478	-8.4	101	0.00
45 C	1,2-Dichloropropane	0.433	0.450	-3.9#	96	0.00
46 T	Dibromomethane	0.239	0.238	0.4	93	0.00
47 T	Bromodichloromethane	0.615	0.629	-2.3	95	0.00
48 T	Methyl methacrylate	0.237	0.259	-9.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051622\
 Data File : VD073031.D
 Acq On : 16 May 2022 12:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00 G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 17 03:27:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	105	0.00
50 S	Toluene-d8	1.260	1.308	-3.8	98	0.00
51 T	4-Methyl-2-Pentanone	0.252	0.271	-7.5	97	0.00
52 CM	Toluene	1.029	1.155	-12.2#	101	0.00
53 T	t-1,3-Dichloropropene	0.564	0.584	-3.5	94	0.00
54 T	cis-1,3-Dichloropropene	0.656	0.691	-5.3	97	0.00
55 T	1,1,2-Trichloroethane	0.323	0.338	-4.6	97	0.00
56 T	Ethyl methacrylate	0.380	0.425	-11.8	96	0.00
57 T	1,3-Dichloropropane	0.554	0.576	-4.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.184	0.198	-7.6	92	0.00
59 T	2-Hexanone	0.169	0.189	-11.8	94	0.00
60 T	Dibromochloromethane	0.407	0.422	-3.7	98	0.00
61 T	1,2-Dibromoethane	0.314	0.328	-4.5	97	0.00
62 S	4-Bromofluorobenzene	0.470	0.478	-1.7	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.366	0.428	-16.9	105	0.00
65 PM	Chlorobenzene	1.156	1.298	-12.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.435	0.503	-15.6	101	0.00
67 C	Ethyl Benzene	2.001	2.426	-21.2#	103	0.00
68 T	m/p-Xylenes	0.771	0.933	-21.0	101	0.00
69 T	o-Xylene	0.720	0.870	-20.8	101	0.00
70 T	Styrene	1.247	1.496	-20.0	100	0.00
71 P	Bromoform	0.236	0.264	-11.9	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.890	4.795	-23.3	107	0.00
74 T	N-amyl acetate	0.969	1.081	-11.6	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.785	0.861	-9.7	102	0.00
76 T	1,2,3-Trichloropropane	0.558	0.595	-6.6	92	0.00
77 T	Bromobenzene	0.924	1.034	-11.9	100	0.00
78 T	n-propylbenzene	4.903	5.963	-21.6	105	0.00
79 T	2-Chlorotoluene	2.864	3.308	-15.5	103	0.00
80 T	1,3,5-Trimethylbenzene	3.299	4.038	-22.4	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.242	0.273	-12.8	97	0.00
82 T	4-Chlorotoluene	2.992	3.452	-15.4	102	0.00
83 T	tert-Butylbenzene	2.804	3.580	-27.7#	110	0.00
84 T	1,2,4-Trimethylbenzene	3.297	3.909	-18.6	102	0.00
85 T	sec-Butylbenzene	4.230	5.342	-26.3#	108	0.00
86 T	p-Isopropyltoluene	3.468	4.354	-25.5#	107	0.00
87 T	1,3-Dichlorobenzene	1.861	2.124	-14.1	104	0.00
88 T	1,4-Dichlorobenzene	1.853	2.090	-12.8	103	0.00
89 T	n-Butylbenzene	3.359	4.226	-25.8#	107	0.00
90 T	Hexachloroethane	0.710	0.852	-20.0	108	0.00
91 T	1,2-Dichlorobenzene	1.622	1.829	-12.8	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.124	0.136	-9.7	100	0.00
93 T	1,2,4-Trichlorobenzene	0.960	1.133	-18.0	102	0.00
94 T	Hexachlorobutadiene	0.525	0.635	-21.0	110	0.00
95 T	Naphthalene	1.766	2.131	-20.7	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051622\
Data File : VD073031.D
Acq On : 16 May 2022 12:48
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00 G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 17 03:27:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.851	0.995	-16.9	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5