

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051723\
 Data File : VD076125.D
 Acq On : 17 May 2023 16:15
 Operator : KP/SY
 Sample : 02811-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-11995

Quant Time: May 18 05:21:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 10:08:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	159352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	306404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	250373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	67897	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	109779	71.281	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.560%
35) Dibromofluoromethane	7.805	113	134980	63.732	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.460%
50) Toluene-d8	10.269	98	377894	45.771	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.540%
62) 4-Bromofluorobenzene	12.575	95	83699	36.495	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	72.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051723\
Data File : VD076125.D
Acq On : 17 May 2023 16:15
Operator : KP/SY
Sample : 02811-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-11995

Quant Time: May 18 05:21:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
Quant Title : SW846 8260
QLast Update : Tue May 16 10:08:53 2023
Response via : Initial Calibration

