


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051723\
 Data File : VD076118.D
 Acq On : 17 May 2023 11:07
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051523S.M
 Title : SW846 8260
 Last Update : Tue May 16 10:08:53 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	13388	PASS
75	95	30	60	49.5	37085	PASS
95	95	100	100	100.0	74883	PASS
96	95	5	9	7.2	5411	PASS
173	174	0.00	2	1.1	766	PASS
174	95	50	100	92.1	68952	PASS
175	174	5	9	7.0	4823	PASS
176	174	95	101	99.2	68427	PASS
177	176	5	9	6.0	4110	PASS