

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073093.D
 Acq On : 18 May 2022 12:25
 Operator : VA/SY
 Sample : N2847-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PSC-124019

Quant Time: May 18 16:29:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	117322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	210857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	235892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	108126	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	99815	68.031	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.060%
35) Dibromofluoromethane	7.909	113	84192	56.869	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.740%
50) Toluene-d8	10.332	98	261982	48.779	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.560%
62) 4-Bromofluorobenzene	12.620	95	100198	51.508	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073093.D
Acq On : 18 May 2022 12:25
Operator : VA/SY
Sample : N2847-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PSC-124019

Quant Time: May 18 16:29:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

