

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073094.D
 Acq On : 18 May 2022 12:53
 Operator : VA/SY
 Sample : N2855-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-051322

Quant Time: May 18 16:29:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	312963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	570694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	572455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	258598	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	171729	43.877	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.760%
35) Dibromofluoromethane	7.908	113	190593	47.566	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.140%
50) Toluene-d8	10.332	98	700644	48.200	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.620	95	251727	47.812	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073094.D
Acq On : 18 May 2022 12:53
Operator : VA/SY
Sample : N2855-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-051322

Quant Time: May 18 16:29:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

