

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073096.D
 Acq On : 18 May 2022 13:50
 Operator : VA/SY
 Sample : N2851-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-051322

Quant Time: May 18 16:30:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	300487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	535858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	544848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	242535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	169682	45.154	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	90.300%		
35) Dibromofluoromethane	7.914	113	179126	47.610	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	95.220%		
50) Toluene-d8	10.332	98	661826	48.489	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	96.980%		
62) 4-Bromofluorobenzene	12.620	95	232250	46.980	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	93.960%		
Target Compounds					Qvalue	
16) Acetone	4.150	43	7132	10.488	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073096.D
Acq On : 18 May 2022 13:50
Operator : VA/SY
Sample : N2851-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-051322

Quant Time: May 18 16:30:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

