

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073097.D
 Acq On : 18 May 2022 14:19
 Operator : VA/SY
 Sample : N2824-02
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COM-2

Quant Time: May 19 04:44:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	301850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	543651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	564831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	262860	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	177661	47.064	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.120%
35) Dibromofluoromethane	7.908	113	111251	29.146	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	58.300%
50) Toluene-d8	10.332	98	672169	48.541	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.080%
62) 4-Bromofluorobenzene	12.620	95	250332	49.912	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.820%
Target Compounds						
16) Acetone	4.156	43	12471	18.256	ug/l	98
95) Naphthalene	15.379	128	72030	10.694	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073097.D
Acq On : 18 May 2022 14:19
Operator : VA/SY
Sample : N2824-02
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COM-2

Quant Time: May 19 04:44:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

