

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073100.D
 Acq On : 18 May 2022 15:45
 Operator : VA/SY
 Sample : N2832-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB01

Quant Time: May 19 04:45:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	330407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	582481	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	590511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	268413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	173072	41.886	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.780%
35) Dibromofluoromethane	7.909	113	193073	47.210	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.420%
50) Toluene-d8	10.332	98	712819	48.045	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.080%
62) 4-Bromofluorobenzene	12.620	95	273092	50.820	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.640%
Target Compounds						
16) Acetone	4.162	43	26979	36.081	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073100.D
Acq On : 18 May 2022 15:45
Operator : VA/SY
Sample : N2832-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB01

Quant Time: May 19 04:45:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

