

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
 Data File : VD073119.D
 Acq On : 19 May 2022 02:18
 Operator : VA/SY
 Sample : N2816-22
 Misc : 5.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P007-SS014-0006-01

Quant Time: May 19 04:55:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	290755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	527755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	538174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	254021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	166249	45.722	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.440%
35) Dibromofluoromethane	7.909	113	175366	47.327	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.660%
50) Toluene-d8	10.332	98	639996	47.610	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.220%
62) 4-Bromofluorobenzene	12.620	95	242285	49.762	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.520%
Target Compounds						
20) Methylene Chloride	4.962	84	28722	6.014	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051822\
Data File : VD073119.D
Acq On : 19 May 2022 02:18
Operator : VA/SY
Sample : N2816-22
Misc : 5.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P007-SS014-0006-01

Quant Time: May 19 04:55:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

