

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
 Data File : VD069193.D
 Acq On : 19 May 2021 13:36
 Operator : VA/SY
 Sample : M2439-03
 Misc : 6.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-282

Quant Time: May 20 01:00:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	210401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	402640	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	365602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	125342	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	146322	69.37	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.74%
35) Dibromofluoromethane	7.920	113	147513	56.30	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.60%
50) Toluene-d8	10.338	98	498263	51.75	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.50%
62) 4-Bromofluorobenzene	12.632	95	145126	46.85	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
Data File : VD069193.D
Acq On : 19 May 2021 13:36
Operator : VA/SY
Sample : M2439-03
Misc : 6.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-282

Quant Time: May 20 01:00:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

