

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
 Data File : VD069199.D
 Acq On : 19 May 2021 16:27
 Operator : VA/SY
 Sample : M2437-07
 Misc : 4.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3

Quant Time: May 20 01:02:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	190303	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	377649	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	368920	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	150910	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	138926	72.82	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.64%
35) Dibromofluoromethane	7.914	113	143215	58.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.54%
50) Toluene-d8	10.338	98	490609	54.33	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.66%
62) 4-Bromofluorobenzene	12.626	95	171906	59.16	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.32%
Target Compounds						
32) 1,1,1-Trichloroethane	7.903	97	47326	12.86	ug/l	99
64) Tetrachloroethene	10.873	164	5785	2.26	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
Data File : VD069199.D
Acq On : 19 May 2021 16:27
Operator : VA/SY
Sample : M2437-07
Misc : 4.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3

Quant Time: May 20 01:02:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

