

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
 Data File : VD069207.D
 Acq On : 19 May 2021 20:17
 Operator : VA/SY
 Sample : M2457-01
 Misc : 4.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4

Quant Time: May 20 01:04:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 13 04:01:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	167129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	339080	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	363178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	153228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	129847	77.50	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.00%
35) Dibromofluoromethane	7.914	113	134936	61.15	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.30%
50) Toluene-d8	10.337	98	442177	54.54	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.08%
62) 4-Bromofluorobenzene	12.626	95	155792	59.71	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051921\
Data File : VD069207.D
Acq On : 19 May 2021 20:17
Operator : VA/SY
Sample : M2457-01
Misc : 4.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4

Quant Time: May 20 01:04:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Quant Title : SW846 8260
QLast Update : Thu May 13 04:01:29 2021
Response via : Initial Calibration

