

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD052019\
 Data File : VD062469.D
 Acq On : 20 May 2019 21:19
 Operator : FY/SY
 Sample : K2871-01
 Misc : 5.86g/5mL/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-04(26.5)

Quant Time: May 21 04:03:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D052019S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 21 03:32:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	527237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	838091	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.38	117	510828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	143081	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	232363	41.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.34%	
35) Dibromofluoromethane	6.95	113	245690	32.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.00%	
50) Toluene-d8	10.42	98	460962	27.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.72%	
62) 4-Bromofluorobenzene	13.56	95	141836	19.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.20%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	3.38	76	69454	5.873	ug/l	98
31) Cyclohexane	6.98	56	34751	6.521	ug/l	99
39) Methylcyclohexane	8.88	83	165168	24.992	ug/l	96
73) Isopropylbenzene	13.41	105	95449	10.583	ug/l	91
78) n-propylbenzene	13.78	91	125855	12.047	ug/l	96
83) tert-Butylbenzene	14.20	119	25353	2.927	ug/l	96
84) 1,2,4-Trimethylbenzene	14.25	105	11460	1.432	ug/l	81
85) sec-Butylbenzene	14.38	105	142395	15.765	ug/l	100
89) n-Butylbenzene	14.81	91	144973	18.353	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD052019\
Data File : VD062469.D
Acq On : 20 May 2019 21:19
Operator : FY/SY
Sample : K2871-01
Misc : 5.86g/5mL/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-04-(26.5)

Quant Time: May 21 04:03:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D052019S.M
Quant Title : SW846 8260
QLast Update : Tue May 21 03:32:39 2019
Response via : Initial Calibration

