



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052021\
Data File : VD069209.D
Acq On : 20 May 2021 09:33
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051221S.M
Title : SW846 8260
Last Update : Thu May 13 04:01:29 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1811

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.7	11952	PASS
75	95	30	60	51.8	37157	PASS
95	95	100	100	100.0	71749	PASS
96	95	5	9	6.5	4698	PASS
173	174	0.00	2	1.5	954	PASS
174	95	50	100	88.8	63731	PASS
175	174	5	9	7.4	4719	PASS
176	174	95	101	97.8	62355	PASS
177	176	5	9	6.6	4090	PASS