

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073193.D
 Acq On : 20 May 2022 19:51
 Operator : VA/SY
 Sample : N2935-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-051822

Quant Time: May 21 00:16:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	317778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	548812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	594076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	243936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	159581	45.440	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.880%
35) Dibromofluoromethane	7.908	113	184646	50.987	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.980%
50) Toluene-d8	10.332	98	736538	55.174	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.340%
62) 4-Bromofluorobenzene	12.620	95	249796	53.320	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.640%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.991	85	3956	1.246	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
Data File : VD073193.D
Acq On : 20 May 2022 19:51
Operator : VA/SY
Sample : N2935-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-051822

Quant Time: May 21 00:16:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

