

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073194.D
 Acq On : 20 May 2022 20:20
 Operator : VA/SY
 Sample : N2938-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-051822

Quant Time: May 21 00:16:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	4978	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	7592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	7631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	3413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	4215	76.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 153.240%	
35) Dibromofluoromethane	7.914	113	3938	78.607	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 157.220%#	
50) Toluene-d8	10.332	98	9767	52.889	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.780%	
62) 4-Bromofluorobenzene	12.620	95	3460	53.388	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
Data File : VD073194.D
Acq On : 20 May 2022 20:20
Operator : VA/SY
Sample : N2938-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-051822

Quant Time: May 21 00:16:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

