

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
 Data File : VD073195.D
 Acq On : 20 May 2022 20:49
 Operator : VA/SY
 Sample : N2942-02
 Misc : 5.14G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RATZER-RD-SP-1-1

Quant Time: May 21 00:17:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	258060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	469453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	504606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	230513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	169205	59.330	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.660%
35) Dibromofluoromethane	7.908	113	169632	54.760	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.520%
50) Toluene-d8	10.332	98	589241	51.601	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.200%
62) 4-Bromofluorobenzene	12.620	95	247520	61.765	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052022\
Data File : VD073195.D
Acq On : 20 May 2022 20:49
Operator : VA/SY
Sample : N2942-02
Misc : 5.14G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RATZER-RD-SP-1-1

Quant Time: May 21 00:17:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

